National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name : TESTING AND CALIBRATION LABORATORY, PRIMA EQUIPMENT, BLOCK NO. 128, DABHASA-UMRAYA ROAD, VADODARA, GUJARAT, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number TC-9355 **Page No** 1 of 1

Validity 25/03/2021 to 24/03/2023 **Last Amended on** -

S.No	Discipline / Group	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used
Permanent Facility				
1	ELECTRONICS-MISCELLANEOUS PRODUCTS	Automated Measuring systems-NO	Influence of sample gas flow for extractive AMS	DIN EN 15267-3 (Clause No. 10.16)
2	ELECTRONICS-MISCELLANEOUS PRODUCTS	Automated Measuring systems-NO	Repeatability standard deviation at span point	DIN EN 15267-3 (Clause No. 10.11)
3	ELECTRONICS-MISCELLANEOUS PRODUCTS	Automated Measuring systems-NO	Repeatability standard deviation at zero point	DIN EN 15267-3 (Clause No. 10.10)
4	ELECTRONICS-MISCELLANEOUS PRODUCTS	Automated Measuring systems-NO	Span Drift	DIN EN 15267-3 (Clause No. 10.13)
5	ELECTRONICS-MISCELLANEOUS PRODUCTS	Automated Measuring systems-NO	Influence of Ambient temperature	DIN EN 15267-3 (Clause No. 10.14)
6	ELECTRONICS-MISCELLANEOUS PRODUCTS	Automated Measuring systems-NO	Influence of Voltage variation	DIN EN 15267-3 (Clause No. 10.17)
7	ELECTRONICS-MISCELLANEOUS PRODUCTS	Automated Measuring systems-NO	Zero drift	DIN EN 15267-3 (Clause No. 10.13)
8	ELECTRONICS-MISCELLANEOUS PRODUCTS	Automated Measuring systems-NO	Lack of fit	DIN EN 15267-3 (Clause No. 10.12)
9	ELECTRONICS-MISCELLANEOUS PRODUCTS	Automated Measuring systems-NO	Response Time	DIN EN 15267-3 (Clause No.10.9)