National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name : WIPRO GE HEALTHCARE PVT. LTD., TESTING LABORATORY, JOHN F WELCH TECHNOLOGY CENTER, # 122, PHASE -5, EPIP-2, BENGALURU, BENGALURU URBAN, KARNATAKA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number TC-10189 **Page No** 1 of 9

Validity 03/01/2022 to 02/01/2024 **Last Amended on** 18/02/2022

S.No	Discipline / Group	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used
Permanent Facility				
1	ELECTRONICS-ENVIRONMENTAL TEST FACILITY	Medica Equipments	Damp heat Steady Test	IEC 60068-2-67
2	ELECTRONICS-ENVIRONMENTAL TEST FACILITY	Medical Equiments	Rapid Change of temperature	IEC 60068-2-14 - (Tests Na & Nb)
3	ELECTRONICS-ENVIRONMENTAL TEST FACILITY	Medical equipment	Classification of ME EQUIPMENT and ME SYSTEMS	IEC/EN 60601-1+A1+A2 (Cl 6)
4	ELECTRONICS-ENVIRONMENTAL TEST FACILITY	Medical Equipment	Vibration - Sinusoidal Testing	IS 9000
5	ELECTRONICS-ENVIRONMENTAL TEST FACILITY	Medical Equipments	Vibration-Random Testing	46-316745_r11 Mechanical Environment Test Guideline of GEHC - (Cl 4.5.4.3, Cl 5.1.2, Cl 6.0, Cl 7.2.1, Cl 7.2.2, Cl 7.2.3)
6	ELECTRONICS-ENVIRONMENTAL TEST FACILITY	Medical Equipments	Bump Testing	IS 9000-7- (Section 2 - Bump)
7	ELECTRONICS-ENVIRONMENTAL TEST FACILITY	Medical Equipments	Bump Testing	JSS 55555 (TEST NUMBER 5)
8	ELECTRONICS-ENVIRONMENTAL TEST FACILITY	Medical Equipments	Cold Testing	IEC 60068-2-1
9	ELECTRONICS-ENVIRONMENTAL TEST FACILITY	Medical Equipments	Cold Testing	IS 9000 - (Part 2 - Section 1 to 4)
10	ELECTRONICS-ENVIRONMENTAL TEST FACILITY	Medical Equipments	Cold Testing	JSS 55555 - (Test number 20)
11	ELECTRONICS-ENVIRONMENTAL TEST FACILITY	Medical Equipments	Cold Testing	MIL-STD-810H - (Section 502.7)
12	ELECTRONICS-ENVIRONMENTAL TEST FACILITY	Medical Equipments	Cold Testing / Low temperature	ISO 16750-4 - (Cl 5.1.1)
13	ELECTRONICS-ENVIRONMENTAL TEST FACILITY	Medical Equipments	Damp heat	2252595PRE - Thermal Environment Test Guideline of GEHC - (Cl 4.1, 4.2, 4.3, Appendix B)

SCOPE OF ACCREDITATION

Laboratory Name : WIPRO GE HEALTHCARE PVT. LTD., TESTING LABORATORY, JOHN F WELCH TECHNOLOGY CENTER, # 122, PHASE -5, EPIP-2, BENGALURU, BENGALURU URBAN, KARNATAKA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number TC-10189 **Page No** 2 of 9

Validity 03/01/2022 to 02/01/2024 **Last Amended on** 18/02/2022

S.No	Discipline / Group	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used
14	ELECTRONICS-ENVIRONMENTAL TEST FACILITY	Medical Equipments	Damp heat	MIL-STD-810H - (Section 507.6)
15	ELECTRONICS-ENVIRONMENTAL TEST FACILITY	Medical Equipments	Damp Heat Cyclic Test	IEC 60068-2-30
16	ELECTRONICS-ENVIRONMENTAL TEST FACILITY	Medical Equipments	Damp heat Steady Test	60068-2-78
17	ELECTRONICS-ENVIRONMENTAL TEST FACILITY	Medical Equipments	Damp Heat steady test	JSS 55555 - (Test Number 10)
18	ELECTRONICS-ENVIRONMENTAL TEST FACILITY	Medical Equipments	Damp Heat (Steady state & Cyclic)	IEC 60068-2-38
19	ELECTRONICS-ENVIRONMENTAL TEST FACILITY	Medical Equipments	Dry / High Temperature	MIL-STD-810H - (Section 501.7)
20	ELECTRONICS-ENVIRONMENTAL TEST FACILITY	Medical Equipments	Dry Heat/ High Temperature	IEC 60068-2-2
21	ELECTRONICS-ENVIRONMENTAL TEST FACILITY	Medical Equipments	Dry Heat/ High Temperature	IS 9000 -Part 3
22	ELECTRONICS-ENVIRONMENTAL TEST FACILITY	Medical Equipments	Dry Heat/ High Temperature	ISO 16750-4 - (Cl 5.1.2)
23	ELECTRONICS-ENVIRONMENTAL TEST FACILITY	Medical Equipments	Dry Heat/ High Temperature	JSS 55555 - (Test number 17)
24	ELECTRONICS-ENVIRONMENTAL TEST FACILITY	Medical Equipments	Rapid temperature cycleing / Thermal shock	JSS - 55555 - (Test Number 22)
25	ELECTRONICS-ENVIRONMENTAL TEST FACILITY	Medical Equipments	Shock Testing	46-316745_r11 Mechanical Environment Test Guideline of GEHC - (Cl 4.5.4.4)
26	ELECTRONICS-ENVIRONMENTAL TEST FACILITY	Medical Equipments	Shock Testing	IEC 60068-2-27
27	ELECTRONICS-ENVIRONMENTAL TEST FACILITY	Medical Equipments	Shock Testing	IS 9000 - (Part 7 Section 1 - Shock)

SCOPE OF ACCREDITATION

Laboratory Name : WIPRO GE HEALTHCARE PVT. LTD., TESTING LABORATORY, JOHN F WELCH TECHNOLOGY CENTER, # 122, PHASE -5, EPIP-2, BENGALURU, BENGALURU URBAN, KARNATAKA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number TC-10189 **Page No** 3 of 9

Validity 03/01/2022 to 02/01/2024 **Last Amended on** 18/02/2022

S.No	Discipline / Group	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used
28	ELECTRONICS-ENVIRONMENTAL TEST FACILITY	Medical Equipments	Shock Testing	JSS 55555 - (TEST NUMBER 24)
29	ELECTRONICS-ENVIRONMENTAL TEST FACILITY	Medical Equipments	Shock Testing	MIL-STD-810H - (Method 516.8)
30	ELECTRONICS-ENVIRONMENTAL TEST FACILITY	Medical Equipments	Temperature cyclic tests	ISO 16750-4 - (Cl 5.3)
31	ELECTRONICS-ENVIRONMENTAL TEST FACILITY	Medical Equipments	Temperature Shock	MIL-STD-810H - (Section 503.7)
32	ELECTRONICS-ENVIRONMENTAL TEST FACILITY	Medical Equipments	Temperature Shock/Rapid Change of temperature	IS 9000 - Part 14
33	ELECTRONICS-ENVIRONMENTAL TEST FACILITY	Medical Equipments	Vibration - Random Testing	IEC 60068-2-64
34	ELECTRONICS-ENVIRONMENTAL TEST FACILITY	Medical Equipments	Vibration - Random Testing	ISO 16750-3
35	ELECTRONICS-ENVIRONMENTAL TEST FACILITY	Medical Equipments	Vibration - Random Testing	JSS 55555 - (Cl 4.2.2.1 & Cl 4.2.2.2 [b])
36	ELECTRONICS-ENVIRONMENTAL TEST FACILITY	Medical equipments	Vibration Sinusoidal Testing	IEC 60068-2-6
37	ELECTRONICS-ENVIRONMENTAL TEST FACILITY	Medical Equipments	Vibration Sinusoidal Testing	JSS-55555 - (Test number - 28)
38	ELECTRONICS-ENVIRONMENTAL TEST FACILITY	Medical Equipments	Vibration-Random Testing	MIL-STD-810H (Method 514.8)
39	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipement's	Spillage on ME EQUIPMENT and ME SYSTEMS	IS 13450 (Part 1 - Cl 11.6.3)
40	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipment	Accidental selection of excessive output values	IEC/EN 60601-1+A1+A2 (Cl 12.4.3)
41	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipment	Accuracy of controls and instruments	IEC/EN 60601-1+A1+A2 (Cl 12.1)

SCOPE OF ACCREDITATION

Laboratory Name : WIPRO GE HEALTHCARE PVT. LTD., TESTING LABORATORY, JOHN F WELCH TECHNOLOGY CENTER, # 122, PHASE -5, EPIP-2, BENGALURU, BENGALURU URBAN, KARNATAKA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number TC-10189 **Page No** 4 of 9

Validity 03/01/2022 to 02/01/2024 **Last Amended on** 18/02/2022

S.No	Discipline / Group	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used
42	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipment	Application/ Type/ Cross-sectional area of POWER SUPPLY CORD conductors	IEC/EN 60601-1+A1+A2 (Cl 8.11.3.1 to Cl 8.11.3.2)
43	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipment	Construction of ME EQUIPMENT	IEC/EN 60601-1+A1+A2 (Cl 15.1 to 15.2)
44	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipment	Construction of transformers used to provide separation as required by 8.5	IEC/EN 60601-1+A1+A2 (Cl 15.5.3)
45	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipment	General requirements for testing ME EQUIPMENT	IEC/EN 60601-1+A1+A2 (Cl 5.1 to 5.6)
46	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipment	Humidity preconditioning treatment	IEC/EN 60601-1+A1+A2 (Cl 5.7)
47	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipment	Incorrect output	IEC/EN 60601-1+A1+A2 (Cl 12.4.4)
48	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipment	Indication relevant to safety	IEC/EN 60601-1+A1+A2 (Cl 12.4.2)
49	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipment	Instability HAZARDS	IEC/EN 60601-1+A1+A2 (Cl 9.4)
50	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipment	Insulation	IEC/EN 60601-1+A1+A2 (Cl 8.8)
51	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipment	Intentional exceeding of safety limits	IEC/EN 60601-1+A1+A2 (Cl 12.4.1)
52	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipment	Internal wiring of ME EQUIPMENT	IEC/EN 60601-1+A1+A2 (Cl 15.4.8)
53	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipment	Interruption of the power supply / SUPPLY MAINS to ME EQUIPMENT	IEC/EN 60601-1+A1+A2 (Cl 11.8)
54	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipment	Isolation from the SUPPLY MAINS	IEC/EN 60601-1+A1+A2 (Cl 8.11.1)
55	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipment	Leakage Current and Patient Auxiliary Currents	IEC/EN 60601-1+A1+A2 (Cl 8.7)
56	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipment	Legibility of Markings	IEC/EN 60601-1+A1+A2 (Cl 7.1.2)
57	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipment	MAINS TERMINAL DEVICES	IEC/EN 60601-1+A1+A2 (Cl 8.11.4)
58	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipment	MECHANICAL HAZARD associated with surfaces, corners and edges	IEC/EN 60601-1+A1+A2 (Cl 9.3)

SCOPE OF ACCREDITATION

Laboratory Name :	WIPRO GE HEALTHCARE PVT. LTD., TESTING LABORATORY, JOHN F WELCH TECHNOLOGY CENTER, # 122, PHASE -5, EPIP-2, BENGALURU, BENGALURU URBAN, KARNATAKA, INDIA		
Accreditation Standard	ISO/IEC 17025:2017		
Certificate Number	TC-10189	**Page No**	5 of 9
Validity	03/01/2022 to 02/01/2024	**Last Amended on**	18/02/2022

S.No	Discipline / Group	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used
59	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipment	MECHANICAL HAZARDS associated with moving parts	IEC/EN 60601-1+A1+A2 (Cl 9.2)
60	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipment	MECHANICAL HAZARDS associated with support systems	IEC/EN 60601-1+A1+A2 (Cl 9.8)
61	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipment	Mechanical strength	IEC/EN 60601-1+A1:2012+A2 - (Cl 15.3.1 to 15.4.6.2)
62	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipment	MULTIPLE SOCKET-OUTLETS	IEC/EN 60601-1+A1+A2 (Cl 8.11.2)
63	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipment	Oil containers	IEC/EN 60601-1+A1+A2 (Cl 15.4.9)
64	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipment	Overflow in ME EQUIPMENT	IEC/EN 60601-1+A1+A2 (Cl 11.6.2)
65	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipment	Overload test	IEC/EN 60601-1+A1+A2 (Cl 15.5.1.3)
66	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipment	Power Input	IEC/EN 60601-1+A1+A2 (Cl 4.11)
67	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipment	Protection against excessive temperatures and other HAZARDS	IEC/EN 60601-1+A1+A2 (Cl 11.1)
68	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipment	Spaces filled by insulating compound	IEC/EN 60601-1+A1+A2 (Cl 8.9.3)
69	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipment	Spillage on ME EQUIPMENT and ME SYSTEMS	IEC/EN 60601-1: +A1+A2 (Cl 11.6.3)
70	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipment	USABILITY of ME EQUIPMENT	IEC/EN 60601-1+A1+A2 (Cl 12.2)
71	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipment	USABILITY of the identification, marking and documents	IEC/EN 60601-1+A1+A2 (Cl 7.1.1)
72	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipment's	Accidental selection of excessive output values	IS 13450 - (Part 1 - 12.4.3)
73	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipment's	ACCOMPANYING DOCUMENTS	IEC/EN 60601-1+A1+A2 (Cl 7.9.1 and 7.9.2.1 to 7.9.2.19)
74	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipment's	Accuracy of controls and instruments	IS 13450 - (Part 1 - Cl 12.1)
75	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipment's	Application/ Type/ Cross-sectional area of POWER SUPPLY CORD conductors	IS 13450 (Part 1) (Cl 8.11.3.1 to Cl 8.11.3.2)
76	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipment's	Arrangement of MAINS TERMINAL DEVICES	IS 13450 - (Part 1 - Cl 8.11.4)

SCOPE OF ACCREDITATION

Laboratory Name : WIPRO GE HEALTHCARE PVT. LTD., TESTING LABORATORY, JOHN F WELCH TECHNOLOGY CENTER, # 122, PHASE -5, EPIP-2, BENGALURU, BENGALURU URBAN, KARNATAKA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number TC-10189 **Page No** 6 of 9

Validity 03/01/2022 to 02/01/2024 **Last Amended on** 18/02/2022

S.No	Discipline / Group	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used
77	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipment's	Classification of ME EQUIPMENT and ME SYSTEMS	IS 13450 - (Part 1 - Cl 6)
78	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipment's	Colors of the insulation of conductors	IEC/EN 60601-1: 2005+A1:2012+A2 (Cl 7.7.1 to 7.7.5)
79	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipment's	Colors of the insulation of conductors	IS 13450 - (Part 1 - Cl 7.7.1 to 7.7.5)
80	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipment's	Connections between different parts of ME EQUIPMENT	IS 13450 - (Part 1 - Cl 8.10)
81	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipment's	Construction of ME EQUIPMENT	IS 13450 - (Part 1 - Cl 15.1 to 15.2)
82	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipment's	Construction of transformers used to provide separation as required by 8.5	IS 13450 (Part 1) (Cl 15.5.3)
83	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipment's	Defibrillation protection	IEC/EN 60601-1+A1+A2 (Cl 8.5.5.1)
84	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipment's	Defibrillation protection	IS 13450- (Part 1- Cl 8.5.5.1)
85	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipment's	Determination of APPLIED PARTS and ACCESSIBLE PARTS	IEC/EN 60601-1+A1+A2 (Cl 5.9.1)
86	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipment's	Durability of Markings	IEC/EN 60601-1+A1A2 (Cl 7.1.3)
87	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipment's	Fundamental rule of protection against electric shock	IEC/EN 60601-1+A1+A2 (Cl 8.1)
88	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipment's	General Requirements	IEC/EN 60601-1+A1+A2 (Cl 4.1 to 4.9)
89	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipment's	General requirements for testing ME EQUIPMENT	IS 13450 -(Part 1 - Cl 5.1 to 5.6)
90	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipment's	Humidity preconditioning treatment	IS 13450 - (Part 1 - Cl 5.7)
91	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipment's	Incorrect output	IS 13450- (Part 1 - Cl 12.4.4)
92	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipment's	Indication relevant to safety	IS 13450 - (Part 1 - Cl 12.4.2)
93	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipment's	Indicator lights and controls	IEC/EN 60601-1: 2005+A1:2012+A2 (Cl 7.8.1 to 7.8.2)
94	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipment's	Instability HAZARDS	IS 13450 - (Part 1 - Cl.9.4)

SCOPE OF ACCREDITATION

Laboratory Name : WIPRO GE HEALTHCARE PVT. LTD., TESTING LABORATORY, JOHN F WELCH TECHNOLOGY CENTER, # 122, PHASE -5, EPIP-2, BENGALURU, BENGALURU URBAN, KARNATAKA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number TC-10189 **Page No** 7 of 9

Validity 03/01/2022 to 02/01/2024 **Last Amended on** 18/02/2022

S.No	Discipline / Group	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used
95	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipment's	Insulation	IS 13450 (Part 1) (Cl 8.8)
96	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipment's	Intentional exceeding of safety limits	IS 13450 (Part 1) (Cl 12.4.1)
97	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipment's	Internal wiring of ME EQUIPMENT	IS 13450 (Part 1) (Cl 15.4.8)
98	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipment's	Interruption of the power supply / SUPPLY MAINS to ME EQUIPMENT	IS 13450 - (Part 1 - Cl 11.8)
99	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipment's	Isolation from the SUPPLY MAINS	IS 13450 - (Part 1 - Cl 8.11.1)
100	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipment's	Legibility of Markings	IS 13450 (Part 1) (Cl 7.1.2)
101	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipment's	Limitation of voltage, current or energy	IEC/EN 60601-1+A1+A2 (Cl 8.4)
102	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipment's	Limitation of voltage, current or energy	IS 13450 (Part 1) (Cl 8.4)
103	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipment's	Marking of controls and instruments	IEC/EN 60601-1:2005+A1:2012+A2 (Cl 7.4.1 to 7.4.3)
104	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipment's	Marking on the inside of ME EQUIPMENT or ME EQUIPMENT parts	C/EN 60601-1+A1+A2 (Cl 7.3.1 to 7.3.8)
105	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipment's	Marking on the outside of ME EQUIPMENT or ME EQUIPMENT part	IEC/EN 60601-1+A1+A2 (Cl 7.2.1 to 7.2.2)
106	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipment's	MECHANICAL HAZARD associated with surfaces, corners and edges	IS 13450 - (Part 1 - Cl 9.3)
107	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipment's	MECHANICAL HAZARDS associated with moving parts	IS 13450 (Part 1) (Cl 9.2)
108	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipment's	MECHANICAL HAZARDS associated with support systems	IS 13450 (Part 1) (Cl 9.8)
109	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipment's	MULTIPLE SOCKET-OUTLETS	IS 13450 (Part 1) (Cl 8.11.2)
110	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipment's	Oil containers	IS 13450 (Part 1) (Cl 15.4.9)
111	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipment's	Overflow in ME EQUIPMENT	IS 13450 (Part 1) (Cl 11.6.2)

SCOPE OF ACCREDITATION

Laboratory Name : WIPRO GE HEALTHCARE PVT. LTD., TESTING LABORATORY, JOHN F WELCH TECHNOLOGY CENTER, # 122, PHASE -5, EPIP-2, BENGALURU, BENGALURU URBAN, KARNATAKA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number TC-10189 **Page No** 8 of 9

Validity 03/01/2022 to 02/01/2024 **Last Amended on** 18/02/2022

S.No	Discipline / Group	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used
112	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipment's	Power Input	IEC/EN 60601-1+A1+A2 (Cl 4.11)
113	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipment's	Power Input	IS 13450 (Part 1) (Cl 4.11)
114	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipment's	Power Input	IS 13450 (Part 1) (Cl 4.11)
115	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipment's	Protective earthing, functional earthing and potential equalization of ME EQUIPMENT	IS 13450 (Part 1) (Cl 8.6):
116	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipment's	Protective earthing, functional earthing and potential equalization of ME EQUIPMENT	IEC/EN 60601-1+A1+A2 (Cl 8.6)
117	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipment's	Requirements related to power sources	IEC/EN 60601-1+A1+A2 (Cl 8.2.1 to 8.2.3)
118	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipment's	Requirements related to power sources	IS 13450 (Part 1) (Cl 8.2.1 to 8.2.2)
119	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipment's	Safety signs	IEC/EN 60601-1+A1+A2 (Cl 7.5)
120	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipment's	Safety signs	IS 13450 (Part 1) (Cl 7.5):
121	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipment's	Separation of part	IEC/EN 60601-1+A1+A2 (Cl 8.5.1 to 8.5.4)
122	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipment's	Separation of part	IS 13450 (Part 1) (Cl 8.5.1 to 8.5.4)
123	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipment's	Symbols	IEC/EN 60601-1+A1+A2 (Cl 7.6.1 to 7.6.3)
124	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipment's	Symbols	IS 13450 (Part 1) (Cl 7.6.1 to 7.6.3)
125	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipment's	Technical description	IEC/EN 60601-1+A1+A2 (Cl 7.9.3.1 to 7.9.3.4)
126	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipment's	Technical description	IS 13450 (Part 1) (Cl 7.9.3.1 to 7.9.3.4):
127	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipment's	USABILITY of ME EQUIPMENT	IS 13450 (Part 1) (Cl 12.2)
128	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipment's	USABILITY of the identification, marking and documents	IS 13450 (Part 1)(Cl 7.1.1)
129	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipments	Accessible Parts	IEC/EN 60601-1+A1+A2 (Cl. 5.9.2.1, 5.9.2.2, 5.9.2.3)

SCOPE OF ACCREDITATION

Laboratory Name :	WIPRO GE HEALTHCARE PVT. LTD., TESTING LABORATORY, JOHN F WELCH TECHNOLOGY CENTER, # 122, PHASE -5, EPIP-2, BENGALURU, BENGALURU URBAN, KARNATAKA, INDIA		
Accreditation Standard	ISO/IEC 17025:2017		
Certificate Number	TC-10189	**Page No**	9 of 9
Validity	03/01/2022 to 02/01/2024	**Last Amended on**	18/02/2022

S.No	Discipline / Group	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used
130	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipments	ACCOMPANYING DOCUMENTS	IS 13450 (Part 1)Cl 7.9.1 and 7.9.2.1 to 7.9.2.19)
131	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipments	Electromagnetic compatibility of ME EQUIPMENT and ME SYSTEMS	IEC/EN 60601-1+A1+A2 (Cl 17)
132	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipments	ME SYSTEMS	IEC/EN 60601-1+A1+A2 (Cl 16)
133	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipments	Mechanical Strength	IS 13450 - (Part 1 - Cl 15.3.1 to 15.4.6.2)
134	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipments	POWER INPUT	IEC/EN 60601-1+A1+A2 (Cl 4.11)
135	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipments	Protection against excessive temperatures and other HAZARDS	IS 13450 (Part 1) (Cl 11.1)
136	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipments	SINGLE FAULT CONDITIONS	IEC/EN 60601-1+A1+A2 (Cl 13.2)
137	ELECTRONICS- SAFETY TESTING FACILITY	Medical Equipments	SINGLE FAULT CONDITIONS	IS 13450 - (Part 1 - Cl 13.2)