National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name : V-ACCESS (INDIA) PRIVATE LIMITED, QC LAB, PLOT NO. 12, SECTOR-5, IMT MANESAR, GURUGRAM, HARYANA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number TC-11233 **Page No** 1 of 2

Validity 21/12/2022 to 20/12/2024 **Last Amended on** -

S.No	Discipline / Group	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used
Permanent Facility				
1	MECHANICAL- TEXTILE MATERIALS	Respiratory Protective Devices - Filtering Half Masks	Breathing Resistance - Inhalation and Exhalation Test	BS EN 149 + A1
2	MECHANICAL- TEXTILE MATERIALS	Respiratory Protective Devices - Filtering Half Masks	Breathing Resistance - Inhalation and Exhalation Test	IS 9473
3	MECHANICAL- TEXTILE MATERIALS	Respiratory Protective Devices - Filtering Half Masks	Carbon Dioxide Content of the Inhalation Air Test	BS EN 149 + A1
4	MECHANICAL- TEXTILE MATERIALS	Respiratory Protective Devices - Filtering Half Masks	Carbon Dioxide Content of the Inhalation Air Test	IS 9473
5	MECHANICAL- TEXTILE MATERIALS	Respiratory Protective Devices - Filtering Half Masks	Cleaning and Disinfecting	BS EN 149 + A1
6	MECHANICAL- TEXTILE MATERIALS	Respiratory Protective Devices - Filtering Half Masks	Cleaning and Disinfecting	IS 9473
7	MECHANICAL- TEXTILE MATERIALS	Respiratory Protective Devices - Filtering Half Masks	Field of Vision Test	BS EN 149 + A1
8	MECHANICAL- TEXTILE MATERIALS	Respiratory Protective Devices - Filtering Half Masks	Flammability Test	BS EN 149 + A1
9	MECHANICAL- TEXTILE MATERIALS	Respiratory Protective Devices - Filtering Half Masks	Flammability Test	IS 9473
10	MECHANICAL- TEXTILE MATERIALS	Respiratory Protective Devices - Filtering Half Masks	Packing and Marking	BS EN 149 + A1
11	MECHANICAL- TEXTILE MATERIALS	Respiratory Protective Devices - Filtering Half Masks	Packing and Marking	IS 9473
12	MECHANICAL- TEXTILE MATERIALS	Respiratory Protective Devices - Filtering Half Masks	Particulate Filter Efficiency N95 Test with Paraffin Oil	BS EN 149 + A1
13	MECHANICAL- TEXTILE MATERIALS	Respiratory Protective Devices - Filtering Half Masks	Particulate Filter Efficiency N95 Test with Paraffin Oil	IS 9473
14	MECHANICAL- TEXTILE MATERIALS	Respiratory Protective Devices - Filtering Half Masks	Particulate Filter Efficiency N95 Test with Sodium Chloride	BS EN 149 + A1
15	MECHANICAL- TEXTILE MATERIALS	Respiratory Protective Devices - Filtering Half Masks	Particulate Filter Efficiency N95 Test with Sodium Chloride	IS 9473
16	MECHANICAL- TEXTILE MATERIALS	Respiratory Protective Devices - Filtering Half Masks	Temperature Conditioning @ (-30 °C and +70 °C)	BS EN 149 + A1
17	MECHANICAL- TEXTILE MATERIALS	Respiratory Protective Devices - Filtering Half Masks	Temperature Conditioning @ (-30 °C and +70 °C)	IS 9473
18	MECHANICAL- TEXTILE MATERIALS	Respiratory Protective Devices - Filtering Half Masks	Total Inward Leakage Test - Face Seal Leakage	BS EN 149 + A1

National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name : V-ACCESS (INDIA) PRIVATE LIMITED, QC LAB, PLOT NO. 12, SECTOR-5, IMT MANESAR, GURUGRAM, HARYANA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number TC-11233 **Page No** 2 of 2

Validity 21/12/2022 to 20/12/2024 **Last Amended on** -

S.No	Discipline / Group	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used
19	MECHANICAL- TEXTILE MATERIALS	Respiratory Protective Devices - Filtering Half Masks	Total Inward Leakage Test - Face Seal Leakage	IS 9473
20	MECHANICAL- TEXTILE MATERIALS	Respiratory Protective Devices - Filtering Half Masks	Visual Inspection Test	BS EN 149 + A1
21	MECHANICAL- TEXTILE MATERIALS	Respiratory Protective Devices - Filtering Half Masks	Visual Inspection Test	IS 9473