National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name : PANACEA MEDICAL TECHNOLOGIES PVT LTD, PLOT #35, 4TH PHASE, KIADB INDUSTRIAL AREA, MALUR TALUK, MALUR, KOLAR, KARNATAKA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number TC-10617 **Page No** 1 of 2

Validity 14/05/2022 to 13/05/2024 **Last Amended on** -

S.No	Discipline / Group	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used
Permanent Facility				
1	DIAGNOSTIC RADIOLOGY QA TESTING-MAMMOGRAPHY	Mammography	Consistency of output radiation (CoV)	AERB/RF-MED/SC-03
2	DIAGNOSTIC RADIOLOGY QA TESTING-MAMMOGRAPHY	Mammography	mAs Loading Station (CoL)	AERB/RF-MED/SC-03
3	DIAGNOSTIC RADIOLOGY QA TESTING-MAMMOGRAPHY	Mammography	Operating Potential	AERB/RF-MED/SC-03
4	DIAGNOSTIC RADIOLOGY QA TESTING-MAMMOGRAPHY	Mammography	Performance of imaging phantom	AERB/RF-MED/SC-03
5	DIAGNOSTIC RADIOLOGY QA TESTING-MAMMOGRAPHY	Mammography	Radiation Leakage level from X Ray tube housing	AERB/RF-MED/SC-03
6	DIAGNOSTIC RADIOLOGY QA TESTING-MAMMOGRAPHY	Mammography	Radiation Protection Survey	AERB/RF-MED/SC-03
7	DIAGNOSTIC RADIOLOGY QA TESTING-MAMMOGRAPHY	Mammography	Time	AERB/RF-MED/SC-03
8	DIAGNOSTIC RADIOLOGY QA TESTING-MAMMOGRAPHY	Mammography	Total Filtration	AERB/RF-MED/SC-03

National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name : PANACEA MEDICAL TECHNOLOGIES PVT LTD, PLOT #35, 4TH PHASE, KIADB INDUSTRIAL AREA, MALUR TALUK, MALUR, KOLAR, KARNATAKA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number TC-10617 **Page No** 2 of 2

Validity 14/05/2022 to 13/05/2024 **Last Amended on** -

S.No	Discipline / Group	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used
Site Facility				
1	DIAGNOSTIC RADIOLOGY QA TESTING-MAMMOGRAPHY	Mammography	Consistency of output radiation (CoV)	AERB/RF-MED/SC-03
2	DIAGNOSTIC RADIOLOGY QA TESTING-MAMMOGRAPHY	Mammography	mAs Loading Station (CoL)	AERB/RF-MED/SC-03
3	DIAGNOSTIC RADIOLOGY QA TESTING-MAMMOGRAPHY	Mammography	Operating Potential	AERB/RF-MED/SC-03
4	DIAGNOSTIC RADIOLOGY QA TESTING-MAMMOGRAPHY	Mammography	Performance of imaging phantom	AERB/RF-MED/SC-03
5	DIAGNOSTIC RADIOLOGY QA TESTING-MAMMOGRAPHY	Mammography	Radiation Leakage level from X Ray tube housing	AERB/RF-MED/SC-03
6	DIAGNOSTIC RADIOLOGY QA TESTING-MAMMOGRAPHY	Mammography	Radiation Protection Survey	AERB/RF-MED/SC-03
7	DIAGNOSTIC RADIOLOGY QA TESTING-MAMMOGRAPHY	Mammography	Time	AERB/RF-MED/SC-03
8	DIAGNOSTIC RADIOLOGY QA TESTING-MAMMOGRAPHY	Mammography	Total Filtration	AERB/RF-MED/SC-03