National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name : RELIABLE TECHNOCARE PVT.LTD. (TESTING DIVISION), RELIABLE HOUSE, PALAZZO GREENS, WING-A, UNIT NO.101-104, VADGAON MAVAL, PUNE, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number TC-10159 **Page No** 1 of 8

Validity 22/12/2021 to 21/12/2023 **Last Amended on** -

S.No	Discipline / Group	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used
Site Facility				
1	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	CT Equipment	Computed Tomography Dose Index (CTDI)	AERB Test Protocol
2	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	CT Equipment	High Contrast Resolution	AERB Test Protocol
3	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	CT Equipment	Low Contrast Resolution	AERB Test Protocol
4	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	CT Equipment	mAs Linearity	AERB Test Protocol
5	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	CT Equipment	Operating Potential	AERB Test Protocol
6	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	CT Equipment	Output Consistency	AERB Test Protocol
7	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	CT Equipment	Radiation Profile width/Slice Thickness	AERB Test Protocol
8	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	CT Equipment	Radiation Protection Survey	AERB Test Protocol
9	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	CT Equipment	Timer	AERB Test Protocol
10	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	CT Equipment	Total Filtration	AERB Test Protocol
11	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	CT Equipment	Tube housing leakage	AERB Test Protocol
12	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	Dental Cone Beam CT	CTDI (CTDI W)	AERB Test Protocol
13	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	Dental Cone Beam CT	High Contrast Sensitivity	AERB Test Protocol

SCOPE OF ACCREDITATION

Laboratory Name : RELIABLE TECHNOCARE PVT.LTD. (TESTING DIVISION), RELIABLE HOUSE, PALAZZO GREENS, WING-A, UNIT NO.101-104, VADGAON MAVAL, PUNE, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number TC-10159 **Page No** 2 of 8

Validity 22/12/2021 to 21/12/2023 **Last Amended on** -

S.No	Discipline / Group	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used
14	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	Dental Cone Beam CT	Low contrast Sensitivity	AERB Test Protocol
15	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	Dental Cone Beam CT	mA/mAs Linearity	AERB Test Protocol
16	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	Dental Cone Beam CT	Operating Potential	AERB Test Protocol
17	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	Dental Cone Beam CT	Radiation protection survey	AERB Test Protocol
18	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	Dental Cone Beam CT	Reproducibility of radiation Output	AERB Test Protocol
19	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	Dental Cone Beam CT	Timer	AERB Test Protocol
20	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	Dental Cone Beam CT	Total Filtration	AERB Test Protocol
21	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	Dental Cone Beam CT	Tube housing Leakage	AERB Test Protocol
22	DIAGNOSTIC RADIOLOGY QA TESTING- INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	C-ARM and O-ARM	mA/mAs loading stations	AERB Test Protocol
23	DIAGNOSTIC RADIOLOGY QA TESTING- INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	C-ARM and O-ARM	Exposure Rate at Table Top	AERB Test Protocol
24	DIAGNOSTIC RADIOLOGY QA TESTING- INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	C-ARM and O-ARM	High Contrast Sensitivity	AERB Test Protocol

SCOPE OF ACCREDITATION

Laboratory Name : RELIABLE TECHNOCARE PVT.LTD. (TESTING DIVISION), RELIABLE HOUSE, PALAZZO GREENS, WING-A, UNIT NO.101-104, VADGAON MAVAL, PUNE, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number TC-10159 **Page No** 3 of 8

Validity 22/12/2021 to 21/12/2023 **Last Amended on** -

S.No	Discipline / Group	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used
25	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	C-ARM and O-ARM	Low contrast Sensitivity	AERB Test Protocol
26	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	C-ARM and O-ARM	Operating Potential	AERB Test Protocol
27	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	C-ARM and O-ARM	Output Consistency	AERB Test Protocol
28	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	C-ARM and O-ARM	Total Filtration	AERB Test Protocol
29	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	C-ARM and O-ARM	Tube housing leakage	AERB Test Protocol
30	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Interventional Radiology Equipment	Exposure rate at table top	AERB Test Protocol
31	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Interventional Radiology Equipment	High Contrast Sensitivity	AERB Test Protocol
32	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Interventional Radiology Equipment	Irradiation time	AERB Test Protocol
33	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Interventional Radiology Equipment	Low contrast Sensitivity	AERB Test Protocol

SCOPE OF ACCREDITATION

Laboratory Name : RELIABLE TECHNOCARE PVT.LTD. (TESTING DIVISION), RELIABLE HOUSE, PALAZZO GREENS, WING-A, UNIT NO.101-104, VADGAON MAVAL, PUNE, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number TC-10159 **Page No** 4 of 8

Validity 22/12/2021 to 21/12/2023 **Last Amended on** -

S.No	Discipline / Group	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used
34	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Interventional Radiology Equipment	mA/mAs loading stations	AERB Test Protocol
35	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Interventional Radiology Equipment	Operating Potential	AERB Test Protocol
36	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Interventional Radiology Equipment	Output Consistency	AERB Test Protocol
37	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Interventional Radiology Equipment	Radiation protection survey	AERB Test Protocol
38	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Interventional Radiology Equipment	Total Filtration	AERB Test Protocol
39	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Interventional Radiology Equipment	Tube housing leakage	AERB Test Protocol
40	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	X-ray & Fluoroscopy	Exposure Rate at Table Top	AERB Test Protocol
41	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	X-ray & Fluoroscopy	Irradiation time	AERB Test Protocol
42	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	X-ray & Fluoroscopy	Low contrast Sensitivity	AERB Test Protocol

SCOPE OF ACCREDITATION

Laboratory Name : RELIABLE TECHNOCARE PVT.LTD. (TESTING DIVISION), RELIABLE HOUSE, PALAZZO GREENS, WING-A, UNIT NO.101-104, VADGAON MAVAL, PUNE, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number TC-10159 **Page No** 5 of 8

Validity 22/12/2021 to 21/12/2023 **Last Amended on** -

S.No	Discipline / Group	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used
43	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	X-ray & Fluoroscopy	mA/mAs loading stations	AERB Test Protocol
44	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	X-ray & Fluoroscopy	Operating Potential	AERB Test Protocol
45	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	X-ray & Fluoroscopy	Radiation output consistancy	AERB Test Protocol
46	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	X-ray & Fluoroscopy	Radiation protection survey	AERB Test Protocol
47	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	X-ray & Fluoroscopy	Total Filtration	AERB Test Protocol
48	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	X-ray & Fluoroscopy	Tube housing leakage	AERB Test Protocol
49	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	X-ray & Flurosocopy	Central beam alignment	AERB Test Protocol
50	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	X-ray & Flurosocopy	Congruence of radiation and optical field	AERB Test Protocol
51	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	X-ray & Flurosocopy	Effective focal spot measurement	AERB Test Protocol

SCOPE OF ACCREDITATION

Laboratory Name : RELIABLE TECHNOCARE PVT.LTD. (TESTING DIVISION), RELIABLE HOUSE, PALAZZO GREENS, WING-A, UNIT NO.101-104, VADGAON MAVAL, PUNE, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number TC-10159 **Page No** 6 of 8

Validity 22/12/2021 to 21/12/2023 **Last Amended on** -

S.No	Discipline / Group	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used
52	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	X-ray & Flurosocopy	High Contrast Sensitivity	AERB Test Protocol
53	DIAGNOSTIC RADIOLOGY QA TESTING-MAMMOGRAPHY	Mammography	Half Value Layer(HVL)	AERB Test Protocol
54	DIAGNOSTIC RADIOLOGY QA TESTING-MAMMOGRAPHY	Mammography	Irradiation time	AERB Test Protocol
55	DIAGNOSTIC RADIOLOGY QA TESTING-MAMMOGRAPHY	Mammography	mA/mAs loading stations	AERB Test Protocol
56	DIAGNOSTIC RADIOLOGY QA TESTING-MAMMOGRAPHY	Mammography	Operating Potential	AERB Test Protocol
57	DIAGNOSTIC RADIOLOGY QA TESTING-MAMMOGRAPHY	Mammography	Performance of Imaging Phantom	AERB Test Protocol
58	DIAGNOSTIC RADIOLOGY QA TESTING-MAMMOGRAPHY	Mammography	Radiation output consistancy	AERB Test Protocol
59	DIAGNOSTIC RADIOLOGY QA TESTING-MAMMOGRAPHY	Mammography	Radiation protection survey	AERB Test Protocol
60	DIAGNOSTIC RADIOLOGY QA TESTING-MAMMOGRAPHY	Mammography	Tube housing leakage	AERB Test Protocol
61	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	Dental (OPG and Intra oral)	mA/mAs loading stations	AERB Test Protocol
62	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	Dental (OPG and Intra oral)	Operating Potential	AERB Test Protocol
63	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	Dental (OPG and Intra oral)	Radiation protection survey	AERB Test Protocol

SCOPE OF ACCREDITATION

Laboratory Name : RELIABLE TECHNOCARE PVT.LTD. (TESTING DIVISION), RELIABLE HOUSE, PALAZZO GREENS, WING-A, UNIT NO.101-104, VADGAON MAVAL, PUNE, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number TC-10159 **Page No** 7 of 8

Validity 22/12/2021 to 21/12/2023 **Last Amended on** -

S.No	Discipline / Group	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used
64	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	Dental (OPG and Intra oral)	Reproducibility of radiation Output	AERB Test Protocol
65	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	Dental (OPG and Intra oral)	Timer	AERB Test Protocol
66	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	Dental (OPG and Intra oral)	Total Filtration	AERB Test Protocol
67	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	Dental (OPG and Intra oral)	Tube housing leakage	AERB Test Protocol
68	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	X-ray - (Fixed)	Radiation protector survey	AERB Test Protocol
69	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	X-ray -(Fixed, Portable, Mobile)	Central beam alignment	AERB Test Protocol
70	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	X-ray -(Fixed, Portable, Mobile)	Congruence of radiation and optical field	AERB Test Protocol
71	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	X-ray -(Fixed, Portable, Mobile)	Effective focal spot measurement	AERB Test Protocol
72	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	X-ray -(Fixed, Portable, Mobile)	Irradiation time	AERB Test Protocol
73	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	X-ray -(Fixed, Portable, Mobile)	mA/mAs loading stations	AERB Test Protocol
74	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	X-ray -(Fixed, Portable, Mobile)	Operating Potential	AERB Test Protocol

SCOPE OF ACCREDITATION

Laboratory Name :	RELIABLE TECHNOCARE PVT.LTD. (TESTING DIVISION), RELIABLE HOUSE, PALAZZO GREENS, WING-A, UNIT NO.101-104, VADGAON MAVAL, PUNE, MAHARASHTRA, INDIA		
Accreditation Standard	ISO/IEC 17025:2017		
Certificate Number	TC-10159	**Page No**	8 of 8
Validity	22/12/2021 to 21/12/2023	**Last Amended on**	-

S.No	Discipline / Group	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used
75	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	X-ray -(Fixed, Portable, Mobile)	Radiation output consistency	AERB Test Protocol
76	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	X-ray -(Fixed, Portable, Mobile)	Total Filtration	AERB Test Protocol
77	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	X-ray -(Fixed, Portable, Mobile)	Tube housing leakage	AERB Test Protocol