National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :	NAGMAN CALIBRATION SERVICES LLP (TESTING), NO.168/7, CHENNAI-BANGALORE NATIONAL HIGHWAY NH-4, CHEMBARAMBAKKAM, CHENNAI, THIRUVALLUR, TAMIL NADU, INDIA		
Accreditation Standard	ISO/IEC 17025:2017		
Certificate Number	TC-11235	**Page No**	1 of 1
Validity	22/12/2022 to 21/12/2024	**Last Amended on**	-

S.No	Discipline / Group	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used
Permanent Facility				
1	FLUID FLOW- LIQUIDS	Control Valve	Capacity Test and Inherent Flow Characteristics (Head Pressure Max.- 3 bar, Ambient Temperature- 30±10 °C, Water Density- 997 to 998 kg/m3, Differential Pressure Max.- 3 bar, Travel Length Max.- 300 mm, Valve Diameter- upto 500 mm)	ANSI/ISA 75.02.01