National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name : SAI IMAGING HEALTHCARE PVT. LTD., OFFICE NO.405/406, SHREEJI BIZZ WORLD, KATHE GALLI SIGNAL, NASHIK, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number TC-10729 **Page No** 1 of 7

Validity 21/06/2022 to 20/06/2024 **Last Amended on** -

S.No	Discipline / Group	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used
Site Facility				
1	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	CT Equipment	High contrast resolution	AERB/RF-MED/SC-03
2	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	CT Equipment	Low Contrast Resolution	AERB/RF-MED/SC-03
3	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	CT Equipment	mA/mAs Linearity	AERB/RF-MED/SC-03
4	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	CT Equipment	Operating Potential (kV)	AERB/RF-MED/SC-03
5	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	CT Equipment	Radiation Dose Index CTDi	AERB/RF-MED/SC-03
6	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	CT Equipment	Radiation Leakage level from X-Ray tube housing	AERB/RF-MED/SC-03
7	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	CT Equipment	Radiation Profile width/ Slice Thickness	AERB/RF-MED/SC-03
8	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	CT Equipment	Radiation Protection Survey	AERB/RF-MED/SC-03
9	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	CT Equipment	Reproducibility of output (CoV)	AERB/RF-MED/SC-03
10	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	CT Equipment	Time(s)	AERB/RF-MED/SC-03
11	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	CT Equipment	Timer Linearity	AERB/RF-MED/SC-03
12	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	CT Equipment	Total Filtration	AERB/RF-MED/SC-03
13	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	Dental Cone Beam CT	mA/mAs Linearity	AERB/RF-MED/SC-03

SCOPE OF ACCREDITATION

Laboratory Name :	SAI IMAGING HEALTHCARE PVT. LTD., OFFICE NO.405/406, SHREEJI BIZZ WORLD, KATHE GALLI SIGNAL, NASHIK, MAHARASHTRA, INDIA		
Accreditation Standard	ISO/IEC 17025:2017		
Certificate Number	TC-10729	**Page No**	2 of 7
Validity	21/06/2022 to 20/06/2024	**Last Amended on**	-

S.No	Discipline / Group	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used
14	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	Dental Cone Beam CT	Operating Potential (kV)	AERB/RF-MED/SC-03
15	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	Dental Cone Beam CT	Radiation Leakage level from X-Ray tube housing	AERB/RF-MED/SC-03
16	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	Dental Cone Beam CT	Radiation Protection Survey	AERB/RF-MED/SC-03
17	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	Dental Cone Beam CT	Reproducibility of output (CoV)	AERB/RF-MED/SC-03
18	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	Dental Cone Beam CT	Time(s)	AERB/RF-MED/SC-03
19	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	Dental Cone Beam CT	Timer Linearity	AERB/RF-MED/SC-03
20	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	Dental Cone Beam CT	Total Filtration	AERB/RF-MED/SC-03
21	DIAGNOSTIC RADIOLOGY QA TESTING- INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	C-ARM and O-ARM	Consistency of radiation output (CoV)	AERB/RF-MED/SC-03
22	DIAGNOSTIC RADIOLOGY QA TESTING- INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	C-ARM and O-ARM	Exposure Rate on Table Top	AERB/RF-MED/SC-03
23	DIAGNOSTIC RADIOLOGY QA TESTING- INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	C-ARM and O-ARM	High Contrast Resolution	AERB/RF-MED/SC-03
24	DIAGNOSTIC RADIOLOGY QA TESTING- INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	C-ARM and O-ARM	Irradiation time (s)	AERB/RF-MED/SC-03

SCOPE OF ACCREDITATION

Laboratory Name :	SAI IMAGING HEALTHCARE PVT. LTD., OFFICE NO.405/406, SHREEJI BIZZ WORLD, KATHE GALLI SIGNAL, NASHIK, MAHARASHTRA, INDIA
Accreditation Standard	ISO/IEC 17025:2017
Certificate Number	TC-10729
Validity	21/06/2022 to 20/06/2024

Last Amended on -

S.No	Discipline / Group	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used
25	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	C-ARM and O-ARM	Linearity of mA/mAs loading stations	AERB/RF-MED/SC-03
26	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	C-ARM and O-ARM	Low Contrast Resolution	AERB/RF-MED/SC-03
27	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	C-ARM and O-ARM	Operating Potential (kV)	AERB/RF-MED/SC-03
28	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	C-ARM and O-ARM	Radiation leakage level at 1m from tube housing	AERB/RF-MED/SC-03
29	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	C-ARM and O-ARM	Radiation Protection Survey	AERB/RF-MED/SC-03
30	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	C-ARM and O-ARM	Total Filtration	AERB/RF-MED/SC-03
31	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Interventional Radiology - Cath Lab	Consistency of radiation output (CoV)	AERB/RF-MED/SC-03
32	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Interventional Radiology - Cath Lab	Exposure Rate on Table Top	AERB/RF-MED/SC-03
33	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Interventional Radiology - Cath Lab	High Contrast Resolution	AERB/RF-MED/SC-03

SCOPE OF ACCREDITATION

Laboratory Name : SAI IMAGING HEALTHCARE PVT. LTD., OFFICE NO.405/406, SHREEJI BIZZ WORLD, KATHE GALLI SIGNAL, NASHIK, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number TC-10729 **Page No** 4 of 7

Validity 21/06/2022 to 20/06/2024 **Last Amended on** -

S.No	Discipline / Group	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used
34	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Interventional Radiology - Cath Lab	Irradiation time (s)	AERB/RF-MED/SC-03
35	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Interventional Radiology - Cath Lab	Linearity of mA/mAs loading stations	AERB/RF-MED/SC-03
36	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Interventional Radiology - Cath Lab	Low Contrast Resolution	AERB/RF-MED/SC-03
37	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Interventional Radiology - Cath Lab	Operating Potential (kV)	AERB/RF-MED/SC-03
38	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Interventional Radiology - Cath Lab	Radiation leakage level at 1m from tube housing	AERB/RF-MED/SC-03
39	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Interventional Radiology - Cath Lab	Radiation Protection Survey	AERB/RF-MED/SC-03
40	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Interventional Radiology - Cath Lab	Timer Linearity	AERB/RF-MED/SC-03
41	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Interventional Radiology - Cath Lab	Total Filtration	AERB/RF-MED/SC-03
42	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	X Ray and Fluoroscopy	Irradiation time (s)	AERB/RF-MED/SC-03

SCOPE OF ACCREDITATION

Laboratory Name : SAI IMAGING HEALTHCARE PVT. LTD., OFFICE NO.405/406, SHREEJI BIZZ WORLD, KATHE GALLI SIGNAL, NASHIK, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number TC-10729 **Page No** 5 of 7

Validity 21/06/2022 to 20/06/2024 **Last Amended on** -

S.No	Discipline / Group	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used
43	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	X Ray and Fluoroscopy	Radiation Protection Survey	AERB/RF-MED/SC-03
44	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	X-ray and Fluoroscopy	Central beam alignment	AERB/RF-MED/SC-03
45	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	X-ray and Fluoroscopy	Congruence of radiation and optical field	AERB/RF-MED/SC-03
46	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	X-ray and Fluoroscopy	Consistency of radiation output (CoV)	AERB/RF-MED/SC-03
47	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	X-ray and Fluoroscopy	Effective focal spot measurement FFD= 60 cm	AERB/RF-MED/SC-03
48	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	X-ray and Fluoroscopy	Exposure Rate in Table Top	AERB/RF-MED/SC-03
49	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	X-ray and Fluoroscopy	High Contrast Resolution	AERB/RF-MED/SC-03
50	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	X-ray and Fluoroscopy	Low Contrast Resolution	AERB/RF-MED/SC-03
51	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	X-ray and Fluoroscopy	mA/mAs Linearity	AERB/RF-MED/SC-03

SCOPE OF ACCREDITATION

Laboratory Name : SAI IMAGING HEALTHCARE PVT. LTD., OFFICE NO.405/406, SHREEJI BIZZ WORLD, KATHE GALLI SIGNAL, NASHIK, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number TC-10729 **Page No** 6 of 7

Validity 21/06/2022 to 20/06/2024 **Last Amended on** -

S.No	Discipline / Group	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used
52	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	X-ray and Fluoroscopy	Operating Potential (kV)	AERB/RF-MED/SC-03
53	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	X-ray and Fluoroscopy	Radiation leakage level at 1m from tube housing	AERB/RF-MED/SC-03
54	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	X-ray and Fluoroscopy	Total Filtration	AERB/RF-MED/SC-03
55	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	Dental X-Ray Equipment (OPG and Intra Oral)	Consistency of Radiation Output (CoV)	AERB/RF-MED/SC-03
56	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	Dental X-Ray Equipment (OPG and Intra Oral)	Irradiation Time	AERB/RF-MED/SC-03
57	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	Dental X-Ray Equipment (OPG and Intra Oral)	mA/mAs Linearity	AERB/RF-MED/SC-03
58	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	Dental X-Ray Equipment (OPG and Intra Oral)	Operating Potential (kV)	AERB/RF-MED/SC-03
59	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	Dental X-Ray Equipment (OPG and Intra Oral)	Radiation leakage level from X-ray tube housing	AERB/RF-MED/SC-03
60	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	Dental X-Ray Equipment (OPG and Intra Oral)	Radiation Protection Survey	AERB/RF-MED/SC-03
61	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	Dental X-Ray Equipment (OPG and Intra Oral)	Timer Linearity	AERB/RF-MED/SC-03

SCOPE OF ACCREDITATION

Laboratory Name : SAI IMAGING HEALTHCARE PVT. LTD., OFFICE NO.405/406, SHREEJI BIZZ WORLD, KATHE GALLI SIGNAL, NASHIK, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number TC-10729 **Page No** 7 of 7

Validity 21/06/2022 to 20/06/2024 **Last Amended on** -

S.No	Discipline / Group	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used
62	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	Dental X-Ray Equipment (OPG and Intra Oral)	Total filtration	AERB/RF-MED/SC-03
63	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	X-Ray (Fixed)	Radiation Protection Survey	AERB/RF-MED/SC-03
64	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	X-Ray (Fixed, Portable and Mobile)	Central beam alignment	AERB/RF-MED/SC-03
65	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	X-Ray (Fixed, Portable and Mobile)	Congruence of radiation and optical field	AERB/RF-MED/SC-03
66	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	X-Ray (Fixed, Portable and Mobile)	Consistency of radiation output (CoV)	AERB/RF-MED/SC-03
67	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	X-Ray (Fixed, Portable and Mobile)	Effective focal spot measurement FFD= 60 cm	AERB/RF-MED/SC-03
68	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	X-Ray (Fixed, Portable and Mobile)	Irradiation Time (s)	AERB/RF-MED/SC-03
69	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	X-Ray (Fixed, Portable and Mobile)	Linearity of mA/mAs loading stations	AERB/RF-MED/SC-03
70	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	X-Ray (Fixed, Portable and Mobile)	Operating Potential (kV)	AERB/RF-MED/SC-03
71	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	X-Ray (Fixed, Portable and Mobile)	Radiation leakage level at 1m from tube housing	AERB/RF-MED/SC-03
72	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	X-Ray (Fixed, Portable and Mobile)	Total Filtration	AERB/RF-MED/SC-03