National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name : Q.A. TEST LAB, RICHA EQUIPMENTS PVT. LTD., ZB 5-6/487, ZULFE BENGAL, DILSHAD GARDEN, DELHI, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number TC-8555 **Page No** 1 of 1

Validity 12/11/2021 to 11/11/2023 **Last Amended on** -

S.No	Discipline / Group	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used
Permanent Facility				
1	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Watt hour Meters (Class 1)	ac High Voltage Test	IS 13779: Cl. No. 12.7.6.3
2	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Watt hour Meters (Class 1)	Insulation Resistance Test	IS 13779:1999 (RA 2014) Cl. No. 12.7.6.4
3	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Watt hour Meters (Class 1)	Test of Meter Constant	IS 13779, Cl.No. 11.6, 12.15
4	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Watt hour Meters (Class 1)	Test of No-Load Condition	IS 13779:Cl. No.11. 4, 12.13
5	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Watt hour Meters (Class 1)	Test on Limits of Error	IS 13779: Cl. No. 11.11
6	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Watt-hour Meters (Class 1)	Test of Repeatability of Error	IS 13779: Cl. No.12.17
7	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Watt-hour Meters (Class 1)	Test of Starting Condition	IS 13779: Cl. No.11.5 & 12.14
8	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Watthour Meters (Class-1)	Test of Power consumption	IS 13779