National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :	COAL TESTING LABORATORY, PATHAKHERA, PATHAKHERA AREA, P.O PATHAKHERA, BETUL, MADHYA PRADESH, INDIA		
Accreditation Standard	ISO/IEC 17025:2017		
Certificate Number	TC-5454	**Page No**	1 of 1
Validity	22/04/2022 to 21/04/2024	**Last Amended on**	-

S.No	Discipline / Group	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used
Permanent Facility				
1	CHEMICAL- SOLID FUELS	COAL	ASH	IS 1350 (Part 1)
2	CHEMICAL- SOLID FUELS	COAL	Eq. MOISTURE	IS 1350 (Part 1) Clause: 6.7.1
3	CHEMICAL- SOLID FUELS	COAL	GROSS CALORIFIC VALUE	ASTM D5865, IS 1350 (Part II)
4	CHEMICAL- SOLID FUELS	COAL	VOLATILE MATTER	IS 1350 (Part 1)