National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :	CENTRAL LABORATARY JAMUNA KOTMA AREA, JAMUNA COLLIERY, ANUPPUR, MADHYA PRADESH, INDIA		
Accreditation Standard	ISO/IEC 17025:2017		
Certificate Number	TC-11336	**Page No**	1 of 1
Validity	19/01/2023 to 18/01/2025	**Last Amended on**	-

S.No	Discipline / Group	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used
Permanent Facility				
1	CHEMICAL- SOLID FUELS	Coal	Air Dry Moisture	IS 1350 (Part i)
2	CHEMICAL- SOLID FUELS	Coal	Ash%	IS 1350 (Part I)
3	CHEMICAL- SOLID FUELS	Coal	Equilibrated Moisture	IS 1350 (Part I)
4	CHEMICAL- SOLID FUELS	Coal	Gross Calorific Value	IS 1350 (Part II)
5	CHEMICAL- SOLID FUELS	Coal	Total Moisture%	IS 1350 (Part I)
6	CHEMICAL- SOLID FUELS	Coal	Volatile Matter%	IS 1350 (Part I)