National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name : HINDUSTAN HEALTH CARE, PP-XII 313, PADANNAPALAM, KANNUR, KERALA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number TC-10822 **Page No** 1 of 9

Validity 14/07/2022 to 13/07/2024 **Last Amended on** -

S.No	Discipline / Group	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used
Site Facility				
1	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	CT Equipment	High contrast resolution	AERB/RF-MED/SC-03
2	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	CT Equipment	Low contrast resolution	AERB/RF-MED/SC-03
3	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	CT Equipment	mA/mAs Linearity	AERB/RF-MED/SC-03
4	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	CT Equipment	Operating Potential (kV)	AERB/RF-MED/SC-03
5	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	CT Equipment	Radiation Dose Index CTDi	AERB/RF-MED/SC-03
6	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	CT Equipment	Radiation Leakage level from X-Ray tube housing	AERB/RF-MED/SC-03
7	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	CT Equipment	Radiation Profile width/ Slice Thickness	AERB/RF-MED/SC-03
8	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	CT Equipment	Radiation Protection Survey	AERB/RF-MED/SC-03
9	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	CT Equipment	Reproducibility of output (CoV)	AERB/RF-MED/SC-03
10	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	CT Equipment	Time (s)	AERB/RF-MED/SC-03
11	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	CT Equipment	Timer Linearity	AERB/RF-MED/SC-03
12	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	CT Equipment	Total Filtration	AERB/RF-MED/SC-03
13	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	Dental Cone Beam CT	mA/mAs Linearity	AERB/RF-MED/SC-03

SCOPE OF ACCREDITATION

Laboratory Name : HINDUSTAN HEALTH CARE, PP-XII 313, PADANNAPALAM, KANNUR, KERALA, INDIA
Accreditation Standard ISO/IEC 17025:2017
Certificate Number TC-10822 **Page No** 2 of 9
Validity 14/07/2022 to 13/07/2024 **Last Amended on** -

S.No	Discipline / Group	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used
14	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	Dental Cone Beam CT	Operating Potential (kV)	AERB/RF-MED/SC-03
15	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	Dental Cone Beam CT	Radiation Leakage level from X-Ray tube housing	AERB/RF-MED/SC-03
16	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	Dental Cone Beam CT	Radiation Protection Survey	AERB/RF-MED/SC-03
17	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	Dental Cone Beam CT	Reproducibility of output (CoV)	AERB/RF-MED/SC-03
18	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	Dental Cone Beam CT	Time (s)	AERB/RF-MED/SC-03
19	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	Dental Cone Beam CT	Timer Linearity	AERB/RF-MED/SC-03
20	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	Dental Cone Beam CT	Total Filtration	AERB/RF-MED/SC-03
21	DIAGNOSTIC RADIOLOGY QA TESTING- INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	C-ARM and O-ARM	Consistency of radiation output (CoV)	AERB/RF-MED/SC-03
22	DIAGNOSTIC RADIOLOGY QA TESTING- INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	C-ARM and O-ARM	Exposure Rate on Table Top	AERB/RF-MED/SC-03
23	DIAGNOSTIC RADIOLOGY QA TESTING- INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	C-ARM and O-ARM	High Contrast Resolution	AERB/RF-MED/SC-03
24	DIAGNOSTIC RADIOLOGY QA TESTING- INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	C-ARM and O-ARM	Irradiation time (s)	AERB/RF-MED/SC-03

SCOPE OF ACCREDITATION

Laboratory Name : HINDUSTAN HEALTH CARE, PP-XII 313, PADANNAPALAM, KANNUR, KERALA, INDIA
Accreditation Standard ISO/IEC 17025:2017
Certificate Number TC-10822 **Page No** 3 of 9
Validity 14/07/2022 to 13/07/2024 **Last Amended on** -

S.No	Discipline / Group	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used
25	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	C-ARM and O-ARM	Linearity of mA/mAs loading stations	AERB/RF-MED/SC-03
26	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	C-ARM and O-ARM	Low Contrast Resolution	AERB/RF-MED/SC-03
27	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	C-ARM and O-ARM	Operating Potential (kV)	AERB/RF-MED/SC-03
28	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	C-ARM and O-ARM	Radiation leakage level at 1m from tube housing	AERB/RF-MED/SC-03
29	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	C-ARM and O-ARM	Radiation Protection Survey	AERB/RF-MED/SC-03
30	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	C-ARM and O-ARM	Total Filtration	AERB/RF-MED/SC-03
31	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Interventional Radiology - Cath Lab	Consistency of radiation output (CoV)	AERB/RF-MED/SC-03
32	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Interventional Radiology - Cath Lab	Exposure Rate in Table Top	AERB/RF-MED/SC-03
33	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Interventional Radiology - Cath Lab	High Contrast Resolution	AERB/RF-MED/SC-03

SCOPE OF ACCREDITATION

Laboratory Name : HINDUSTAN HEALTH CARE, PP-XII 313, PADANNAPALAM, KANNUR, KERALA, INDIA
Accreditation Standard ISO/IEC 17025:2017
Certificate Number TC-10822 **Page No** 4 of 9
Validity 14/07/2022 to 13/07/2024 **Last Amended on** -

S.No	Discipline / Group	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used
34	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Interventional Radiology - Cath Lab	Irradiation Time(s)	AERB/RF-MED/SC-03
35	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Interventional Radiology - Cath Lab	Linearity of mA/mAs loading stations	AERB/RF-MED/SC-03
36	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Interventional Radiology - Cath Lab	Low Contrast Resolution	AERB/RF-MED/SC-03
37	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Interventional Radiology - Cath Lab	Operating Potential (kV)	AERB/RF-MED/SC-03
38	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Interventional Radiology - Cath Lab	Radiation leakage level at 1m from tube	AERB/RF-MED/SC-03
39	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Interventional Radiology - Cath Lab	Radiation Protection Survey	AERB/RF-MED/SC-03
40	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Interventional Radiology - Cath Lab	Timer Linearity	AERB/RF-MED/SC-03
41	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Interventional Radiology - Cath Lab	Total Filtration	AERB/RF-MED/SC-03
42	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	X-Ray and Fluoroscopy	Central Beam alignment	AERB/RF-MED/SC-03

SCOPE OF ACCREDITATION

Laboratory Name : HINDUSTAN HEALTH CARE, PP-XII 313, PADANNAPALAM, KANNUR, KERALA, INDIA
Accreditation Standard ISO/IEC 17025:2017
Certificate Number TC-10822 **Page No** 5 of 9
Validity 14/07/2022 to 13/07/2024 **Last Amended on** -

S.No	Discipline / Group	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used
43	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	X-Ray and Fluoroscopy	Congruence of Radiation and Optical field	AERB/RF-MED/SC-03
44	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	X-Ray and Fluoroscopy	Consistency of radiation output (CoV)	AERB/RF-MED/SC-03
45	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	X-Ray and Fluoroscopy	Effective focal spot measurement FFD= 60 cm	AERB/RF-MED/SC-03
46	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	X-Ray and Fluoroscopy	Exposure Rate in Table Top	AERB/RF-MED/SC-03
47	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	X-Ray and Fluoroscopy	High Contrast Resolution	AERB/RF-MED/SC-03
48	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	X-Ray and Fluoroscopy	Irradiation Time (s)	AERB/RF-MED/SC-03
49	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	X-Ray and Fluoroscopy	Low Contrast Resolution	AERB/RF-MED/SC-03
50	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	X-Ray and Fluoroscopy	mA/mAs Linearity	AERB/RF-MED/SC-03
51	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	X-Ray and Fluoroscopy	Operating Potential (kV)	AERB/RF-MED/SC-03

SCOPE OF ACCREDITATION

Laboratory Name : HINDUSTAN HEALTH CARE, PP-XII 313, PADANNAPALAM, KANNUR, KERALA, INDIA
Accreditation Standard ISO/IEC 17025:2017
Certificate Number TC-10822 **Page No** 6 of 9
Validity 14/07/2022 to 13/07/2024 **Last Amended on** -

S.No	Discipline / Group	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used
52	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	X-Ray and Fluoroscopy	Radiation leakage level at 1m from tube housing	AERB/RF-MED/SC-03
53	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	X-Ray and Fluoroscopy	Radiation Protection Survey	AERB/RF-MED/SC-03
54	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	X-Ray and Fluoroscopy	Total Filtration	AERB/RF-MED/SC-03
55	DIAGNOSTIC RADIOLOGY QA TESTING-MAMMOGRAPHY	Mammography	Half Value Layer	AERB/RF-MED/SC-03
56	DIAGNOSTIC RADIOLOGY QA TESTING-MAMMOGRAPHY	Mammography	Linearity of mA/mAs loading stations	AERB/RF-MED/SC-03
57	DIAGNOSTIC RADIOLOGY QA TESTING-MAMMOGRAPHY	Mammography	Operating Potential (kV)	AERB/RF-MED/SC-03
58	DIAGNOSTIC RADIOLOGY QA TESTING-MAMMOGRAPHY	Mammography	Performance of Imaging Phantom	AERB/RF-MED/SC-03
59	DIAGNOSTIC RADIOLOGY QA TESTING-MAMMOGRAPHY	Mammography	Radiation Leakage level from X Ray tube housing	AERB/RF-MED/SC-03
60	DIAGNOSTIC RADIOLOGY QA TESTING-MAMMOGRAPHY	Mammography	Radiation Protection Survey	AERB/RF-MED/SC-03
61	DIAGNOSTIC RADIOLOGY QA TESTING-MAMMOGRAPHY	Mammography	Reproducibility of output (CoV)	AERB/RF-MED/SC-03
62	DIAGNOSTIC RADIOLOGY QA TESTING-MAMMOGRAPHY	Mammography	Time	AERB/RF-MED/SC-03
63	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	Dental X-Ray Equipment (OPG and Intra Oral)	Consistency of Radiation Output (CoV)	AERB/RF-MED/SC-03

SCOPE OF ACCREDITATION

Laboratory Name : HINDUSTAN HEALTH CARE, PP-XII 313, PADANNAPALAM, KANNUR, KERALA, INDIA
Accreditation Standard ISO/IEC 17025:2017
Certificate Number TC-10822 **Page No** 7 of 9
Validity 14/07/2022 to 13/07/2024 **Last Amended on** -

S.No	Discipline / Group	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used
64	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	Dental X-Ray Equipment (OPG and Intra Oral)	Irradiation Time	AERB/RF-MED/SC-03
65	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	Dental X-Ray Equipment (OPG and Intra Oral)	mA/mAs Linearity	AERB/RF-MED/SC-03
66	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	Dental X-Ray Equipment (OPG and Intra Oral)	Operating Potential (kV)	AERB/RF-MED/SC-03
67	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	Dental X-Ray Equipment (OPG and Intra Oral)	Radiation leakage level from X-ray tube housing	AERB/RF-MED/SC-03
68	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	Dental X-Ray Equipment (OPG and Intra Oral)	Radiation Protection Survey	AERB/RF-MED/SC-03
69	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	Dental X-Ray Equipment (OPG and Intra Oral)	Timer linearity (Coefficient of Linearity)	AERB/RF-MED/SC-03
70	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	Dental X-Ray Equipment (OPG and Intra Oral)	Total filtration	AERB/RF-MED/SC-03
71	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	DEXA (Dualenergy X-ray Absorptiometry)	Central Beam Alignment	AERB/RF-MED/SC-03
72	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	DEXA (Dualenergy X-ray Absorptiometry)	Consistency of radiation output (CoV)	AERB/RF-MED/SC-03
73	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	DEXA (Dualenergy X-ray Absorptiometry)	Effective focal spot measurement FFD= 60 cm	AERB/RF-MED/SC-03
74	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	DEXA (Dualenergy X-ray Absorptiometry)	Irradiation TIme (s)	AERB/RF-MED/SC-03

SCOPE OF ACCREDITATION

Laboratory Name : HINDUSTAN HEALTH CARE, PP-XII 313, PADANNAPALAM, KANNUR, KERALA, INDIA
Accreditation Standard ISO/IEC 17025:2017
Certificate Number TC-10822 **Page No** 8 of 9
Validity 14/07/2022 to 13/07/2024 **Last Amended on** -

S.No	Discipline / Group	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used
75	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	DEXA (Dualenergy X-ray Absorptiometry)	Linearity of mA/mAs loading stations	AERB/RF-MED/SC-03
76	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	DEXA (Dualenergy X-ray Absorptiometry)	Operating Potential (kV)	AERB/RF-MED/SC-03
77	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	DEXA (Dualenergy X-ray Absorptiometry)	Radiation leakage level at 1m from tube housing	AERB/RF-MED/SC-03
78	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	DEXA (Dualenergy X-ray Absorptiometry)	Total Filtration	AERB/RF-MED/SC-03
79	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	X-Ray (Fixed)	Radiation Protection Survey	AERB/RF-MED/SC-03
80	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	X-Ray (Fixed, Portable and Mobile)	Central Beam Alignment	AERB/RF-MED/SC-03
81	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	X-Ray (Fixed, Portable and Mobile)	Congruence of Radiation and Optical field	AERB/RF-MED/SC-03
82	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	X-Ray (Fixed, Portable and Mobile)	Consistency of radiation output (CoV)	AERB/RF-MED/SC-03
83	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	X-Ray (Fixed, Portable and Mobile)	Effective focal spot measurement FFD= 60 cm	AERB/RF-MED/SC-03
84	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	X-Ray (Fixed, Portable and Mobile)	Irradiation Time (s)	AERB/RF-MED/SC-03
85	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	X-Ray (Fixed, Portable and Mobile)	Linearity of mA/mAs loading stations	AERB/RF-MED/SC-03

SCOPE OF ACCREDITATION

Laboratory Name : HINDUSTAN HEALTH CARE, PP-XII 313, PADANNAPALAM, KANNUR, KERALA, INDIA
Accreditation Standard ISO/IEC 17025:2017
Certificate Number TC-10822 **Page No** 9 of 9
Validity 14/07/2022 to 13/07/2024 **Last Amended on** -

S.No	Discipline / Group	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used
86	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	X-Ray (Fixed, Portable and Mobile)	Operating Potential (kV)	AERB/RF-MED/SC-03
87	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	X-Ray (Fixed, Portable and Mobile)	Radiation leakage level at 1m from tube housing	AERB/RF-MED/SC-03
88	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	X-Ray (Fixed, Portable and Mobile)	Total Filtration	AERB/RF-MED/SC-03