National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name : GENUS METER TESTING LABORATORY, SPL-3, RIICO INDUSTRIAL AREA, SITAPURA,TONK ROAD, JAIPUR, RAJASTHAN, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number TC-6039 **Page No** 1 of 10

Validity 02/07/2022 to 01/07/2024 **Last Amended on** 01/09/2022

S.No	Discipline / Group	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used
Permanent Facility				
1	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 0.2S, 0.5S)	AC High Voltage test	IS 14697
2	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 0.2S, 0.5S)	AC High Voltage Test	IEC 62053-22
3	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 0.2S, 0.5S)	AC High Voltage Test	IEC 62053-23
4	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 0.2S, 0.5S)	AC High Voltage Test	IS 16444 (Part 2)
5	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 0.2S, 0.5S)	Insulation Resistance Test	IEC 62053-22
6	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 0.2S, 0.5S)	Insulation Resistance Test	IEC 62053-23
7	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 0.2S, 0.5S)	Insulation Resistance Test	IS 14697
8	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 0.2S, 0.5S)	Insulation Resistance Test	IS 16444 (part 2)
9	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 0.2S, 0.5S)	Test of Limits of error	IEC 62053-22
10	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 0.2S, 0.5S)	Test of Limits of error	IEC 62053-23

SCOPE OF ACCREDITATION

Laboratory Name :	GENUS METER TESTING LABORATORY, SPL-3, RIICO INDUSTRIAL AREA, SITAPURA,TONK ROAD, JAIPUR, RAJASTHAN, INDIA		
Accreditation Standard	ISO/IEC 17025:2017		
Certificate Number	TC-6039	**Page No**	2 of 10
Validity	02/07/2022 to 01/07/2024	**Last Amended on**	01/09/2022

S.No	Discipline / Group	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used
11	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 0.2S, 0.5S)	Test of Limits of error	IS 14697
12	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 0.2S, 0.5S)	Test of Limits of error	IS 16444 (Part 2)
13	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 0.2S, 0.5S)	Test of Meter Constant	IEC 62053-22
14	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 0.2S, 0.5S)	Test of Meter Constant	IEC 62053-23
15	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 0.2S, 0.5S)	Test of Meter Constant	IS 14697
16	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 0.2S, 0.5S)	Test of Meter Constant	IS 16444 (Part 2)
17	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 0.2S, 0.5S)	Test of No Load condition	IEC 62053-22
18	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 0.2S, 0.5S)	Test of No load condition	IEC 62053-23
19	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 0.2S, 0.5S)	Test of No Load Condition	IS 14697
20	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 0.2S, 0.5S)	Test of No Load Condition	IS 16444 (Part 2)
21	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 0.2S, 0.5S)	Test of power consumption (In Current circuit)	IEC 62053-22

SCOPE OF ACCREDITATION

Laboratory Name :	GENUS METER TESTING LABORATORY, SPL-3, RIICO INDUSTRIAL AREA, SITAPURA,TONK ROAD, JAIPUR, RAJASTHAN, INDIA		
Accreditation Standard	ISO/IEC 17025:2017		
Certificate Number	TC-6039	**Page No**	3 of 10
Validity	02/07/2022 to 01/07/2024	**Last Amended on**	01/09/2022

S.No	Discipline / Group	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used
22	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 0.2S, 0.5S)	Test of power consumption (In Current circuit)	IEC 62053-23
23	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 0.2S, 0.5S)	Test of power consumption (In Current circuit)	IS 14697
24	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 0.2S, 0.5S)	Test of power consumption (In Current circuit)	IS 16444 (Part 2)
25	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 0.2S, 0.5S)	Test of power consumption (In Voltage circuit)	IEC 62053-22
26	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 0.2S, 0.5S)	Test of power consumption (In Voltage circuit)	IEC 62053-22
27	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 0.2S, 0.5S)	Test of power consumption (In Voltage circuit)	IEC 62053-23
28	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 0.2S, 0.5S)	Test of power consumption (In Voltage circuit)	IEC 62053-23
29	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 0.2S, 0.5S)	Test of power consumption (In Voltage circuit)	IS 14697
30	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 0.2S, 0.5S)	Test of power consumption (In Voltage circuit)	IS 14697
31	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 0.2S, 0.5S)	Test of power consumption (In Voltage circuit)	IS 16444 (Part 2)
32	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 0.2S, 0.5S)	Test of power consumption (In Voltage circuit)	IS 16444 (Part 2)

SCOPE OF ACCREDITATION

Laboratory Name :	GENUS METER TESTING LABORATORY, SPL-3, RIICO INDUSTRIAL AREA, SITAPURA,TONK ROAD, JAIPUR, RAJASTHAN, INDIA		
Accreditation Standard	ISO/IEC 17025:2017		
Certificate Number	TC-6039	**Page No**	4 of 10
Validity	02/07/2022 to 01/07/2024	**Last Amended on**	01/09/2022

S.No	Discipline / Group	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used
33	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 0.2S, 0.5S)	Test of repeatability of error	IEC 62053-22
34	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 0.2S, 0.5S)	Test of repeatability of error	IEC 62053-23
35	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 0.2S, 0.5S)	Test of repeatability of error	IS 14697
36	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 0.2S, 0.5S)	Test of repeatability of error	IS 14697
37	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 0.2S, 0.5S)	Test of repeatability of error	IS 16444 (Part 2)
38	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 0.2S, 0.5S)	Test of Starting Current	IEC 62053-22
39	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 0.2S, 0.5S)	Test of Starting Current	IEC 62053-23
40	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 0.2S, 0.5S)	Test of Starting Current	IS 14697
41	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 0.2S, 0.5S)	Test of Starting current	IS 16444 (Part 2)
42	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 0.2S, 0.5S, 1.0)	AC High Voltage Test	CBIP325
43	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 0.2S, 0.5S, 1.0)	AC High Voltage Test	IEC 62052-11

SCOPE OF ACCREDITATION

Laboratory Name :	GENUS METER TESTING LABORATORY, SPL-3, RIICO INDUSTRIAL AREA, SITAPURA,TONK ROAD, JAIPUR, RAJASTHAN, INDIA		
Accreditation Standard	ISO/IEC 17025:2017		
Certificate Number	TC-6039	**Page No**	5 of 10
Validity	02/07/2022 to 01/07/2024	**Last Amended on**	01/09/2022

S.No	Discipline / Group	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used
44	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 0.2S, 0.5S, 1.0)	Insulation Resistance Test	CBIP-325
45	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 0.2S, 0.5S, 1.0)	Insulation Resistance Test	IEC 62052-11
46	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 0.2S, 0.5S, 1.0)	Test of Limits of error	CBIP-325
47	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 0.2S, 0.5S, 1.0)	Test of Limits of error	IEC 62052-11
48	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 0.2S, 0.5S, 1.0)	Test of meter constant	CBIP-325
49	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 0.2S, 0.5S, 1.0)	Test of Meter constant	IEC 62052-11
50	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 0.2S, 0.5S, 1.0)	Test of No load condition	CBIP-325
51	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 0.2S, 0.5S, 1.0)	Test of No Load Condition	IEC 62052-11
52	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 0.2S, 0.5S, 1.0)	Test of power consumption (In Current circuit)	CBIP 325
53	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 0.2S, 0.5S, 1.0)	Test of power consumption (In Current circuit)	IEC 62052-11
54	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 0.2S, 0.5S, 1.0)	Test of power consumption (In Voltage circuit)	CBIP 325

SCOPE OF ACCREDITATION

Laboratory Name :	GENUS METER TESTING LABORATORY, SPL-3, RIICO INDUSTRIAL AREA, SITAPURA,TONK ROAD, JAIPUR, RAJASTHAN, INDIA		
Accreditation Standard	ISO/IEC 17025:2017		
Certificate Number	TC-6039	**Page No**	6 of 10
Validity	02/07/2022 to 01/07/2024	**Last Amended on**	01/09/2022

S.No	Discipline / Group	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used
55	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 0.2S, 0.5S, 1.0)	Test of power consumption (In Voltage circuit)	CBIP 325
56	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 0.2S, 0.5S, 1.0)	Test of power consumption (In Voltage circuit)	IEC 62052-11
57	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 0.2S, 0.5S, 1.0)	Test of power consumption (In Voltage circuit)	IEC 62052-11
58	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 0.2S, 0.5S, 1.0)	Test of repeatability of error	CBIP-325
59	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 0.2S, 0.5S, 1.0)	Test of repeatability of error	IEC 62052-11
60	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 0.2S, 0.5S, 1.0)	Test of Starting Current	CBIP-325
61	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 0.2S, 0.5S, 1.0)	Test of Starting Current	IEC 62052-11
62	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 1.0)	Test of repeatability of error	IS 16444 (Part 1)
63	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 1.0)	AC High Voltage Test	IEC 62053-21
64	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 1.0)	AC High Voltage Test	IS 13779
65	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 1.0)	AC High Voltage Test	IS 15884

SCOPE OF ACCREDITATION

Laboratory Name :	GENUS METER TESTING LABORATORY, SPL-3, RIICO INDUSTRIAL AREA, SITAPURA,TONK ROAD, JAIPUR, RAJASTHAN, INDIA		
Accreditation Standard	ISO/IEC 17025:2017		
Certificate Number	TC-6039	**Page No**	7 of 10
Validity	02/07/2022 to 01/07/2024	**Last Amended on**	01/09/2022

S.No	Discipline / Group	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used
66	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 1.0)	AC High Voltage Test	IS 16444 (Part 1)
67	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 1.0)	Insulation Resistance Test	IEC 62053-21
68	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 1.0)	Insulation Resistance Test	IS 13779
69	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 1.0)	Insulation Resistance Test	IS 15884
70	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 1.0)	Insulation Resistance Test	IS 16444 (Part 1)
71	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 1.0)	Test of limits of error	IEC 62053-21
72	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 1.0)	Test of Limits of error	IS 13779
73	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 1.0)	Test of Limits of error	IS 15884
74	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 1.0)	Test of Limits of error	IS 16444 (Part 1)
75	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 1.0)	Test of Meter Constant	IEC 62053-21
76	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 1.0)	Test of Meter Constant	IS 13779

SCOPE OF ACCREDITATION

Laboratory Name :	GENUS METER TESTING LABORATORY, SPL-3, RIICO INDUSTRIAL AREA, SITAPURA,TONK ROAD, JAIPUR, RAJASTHAN, INDIA		
Accreditation Standard	ISO/IEC 17025:2017		
Certificate Number	TC-6039	**Page No**	8 of 10
Validity	02/07/2022 to 01/07/2024	**Last Amended on**	01/09/2022

S.No	Discipline / Group	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used
77	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 1.0)	Test of meter constant	IS 13779
78	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 1.0)	Test of meter constant	IS 15884
79	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 1.0)	Test of Meter Constant	IS 16444 (Part 1)
80	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 1.0)	Test of No load condition	IEC 62053-21
81	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 1.0)	Test of No load condition	IS 13779
82	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 1.0)	Test of No Load Condition	IS 15884
83	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 1.0)	Test of No Load Condition	IS 16444 (Part 1)
84	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 1.0)	Test of power consumption (In Current circuit)	IEC 62053-21
85	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 1.0)	Test of power consumption (In Current circuit)	IS 13779
86	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 1.0)	Test of power consumption (In Current circuit)	IS 15884
87	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 1.0)	Test of power consumption (In Current circuit)	IS 16444 (Part 1)

SCOPE OF ACCREDITATION

Laboratory Name :	GENUS METER TESTING LABORATORY, SPL-3, RIICO INDUSTRIAL AREA, SITAPURA,TONK ROAD, JAIPUR, RAJASTHAN, INDIA		
Accreditation Standard	ISO/IEC 17025:2017		
Certificate Number	TC-6039	**Page No**	9 of 10
Validity	02/07/2022 to 01/07/2024	**Last Amended on**	01/09/2022

S.No	Discipline / Group	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used
88	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 1.0)	Test of power consumption (In Voltage circuit)	IEC 62053-21
89	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 1.0)	Test of power consumption (In Voltage circuit)	IEC 62053-21
90	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 1.0)	Test of power consumption (In Voltage circuit)	IS 13779
91	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 1.0)	Test of power consumption (In Voltage circuit)	IS 13779
92	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 1.0)	Test of power consumption (In Voltage circuit)	IS 15884
93	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 1.0)	Test of power consumption (In Voltage circuit)	IS 15884
94	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 1.0)	Test of power consumption (In Voltage circuit)	IS 16444 (Part 1)
95	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 1.0)	Test of power consumption (In Voltage circuit)	IS 16444 (Part 1)
96	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 1.0)	Test of repeatability of error	IEC 62053-21
97	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 1.0)	Test of repeatability of error	IS 13779
98	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 1.0)	Test of repeatability of error	IS 15884

SCOPE OF ACCREDITATION

Laboratory Name :	GENUS METER TESTING LABORATORY, SPL-3, RIICO INDUSTRIAL AREA, SITAPURA,TONK ROAD, JAIPUR, RAJASTHAN, INDIA		
Accreditation Standard	ISO/IEC 17025:2017		
Certificate Number	TC-6039	**Page No**	10 of 10
Validity	02/07/2022 to 01/07/2024	**Last Amended on**	01/09/2022

S.No	Discipline / Group	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used
99	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 1.0)	Test of Starting Current	IEC 62053-21
100	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 1.0)	Test of Starting Current	IS 13779
101	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 1.0)	Test of Starting Current	IS 15884
102	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 1.0)	Test of Starting current	IS 16444 (Part 1)
103	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class: 1.0)	Test on Limits of error	IS 13779
104	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	Electrical & Electronics (Static) Energy Meters (Class:1.0)	Insulation Resistance Test	IS 13779