National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name : METER TESTING LABORATORY, TORRENT POWER LIMITED, PATEL BUILDING, GROUND FLOOR, OSWAL WADI, OLD AGRA ROAD, BHIWANDI, THANE, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number TC-10585 **Page No** 1 of 1

Validity 06/05/2022 to 05/05/2024 **Last Amended on** -

S.No	Discipline / Group	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used
Permanent Facility				
1	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	AC Static Energy Meter 0.2 S, 0.5 S & 1.0 S	Limits of Error	IS 14697
2	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	AC Static Energy Meter 0.2 S, 0.5 S & 1.0 S	Meter Constant	IS 14697
3	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	AC Static Energy Meter 0.2 S, 0.5 S & 1.0 S	No Load Test	IS 14697
4	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	AC Static Energy Meter 0.2 S, 0.5 S & 1.0 S	Starting Current	IS 14697
5	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	AC Static Energy Meter 0.2 S, 0.5 S & 1.0 S	Test of Repeatability	IS 14697
6	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	AC Static Energy Meter Class 1 & 2	Limits of Error	IS 13779
7	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	AC Static Energy Meter Class 1 & 2	Meter Constant	IS 13779
8	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	AC Static Energy Meter Class 1 & 2	No Load Test	IS 13779
9	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	AC Static Energy Meter Class 1 & 2	Repeatability of Error	IS 13779
10	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	AC Static Energy Meter Class 1 & 2	Starting Current	IS 13779