National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name : METER MANAGEMENT GROUP LABORATORY, TATA POWER DELHI DISTRIBUTION LIMITED, LAWRENCE ROAD, KESHAVPURAM, NEW DELHI, NORTH WEST, DELHI, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number TC-5660 **Page No** 1 of 1

Validity 02/09/2022 to 01/09/2024 **Last Amended on** -

S.No	Discipline / Group	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used
Permanent Facility				
1	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	AC STATIC TRNSFORMER OPERATED WATT HOUR & VAR HOUR METERS (Classes - 0.2s & 0.5s)	Limits of Error (Due to variation of the Current)	IS : 14697
2	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	AC STATIC TRNSFORMER OPERATED WATT HOUR & VAR HOUR METERS (Classes - 0.2s & 0.5s)	No load Test	IS : 14697
3	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	AC STATIC TRNSFORMER OPERATED WATT HOUR & VAR HOUR METERS (Classes - 0.2s & 0.5s)	Starting Current	IS : 14697
4	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	AC STATIC WATT HOUR METER (Classes -1 & 2)	Limits of Error (Due to variation of the Current)	IS : 13779
5	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	AC STATIC WATT HOUR METER (Classes -1 & 2)	No load Test	IS : 13779
6	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	AC STATIC WATT HOUR METER (Classes -1 & 2)	Starting Current	IS : 13779
7	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	AC STATIC WATT HOUR SMART METER (Classes - 0.2s, 0.5s & 1.0)	Limits of Error (Due to variation of the Current)	IS : 16444 Part-1& Part-2
8	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	AC STATIC WATT HOUR SMART METER (Classes - 0.2s, 0.5s & 1.0)	No load Test	IS : 16444 Part-1 & Part-2
9	ELECTRICAL- ELECTRICAL INDICATING & RECORDING INSTRUMENTS	AC STATIC WATT HOUR SMART METER (Classes - 0.2s, 0.5s & 1.0)	Starting Current	IS : 16444 Part-1 & Part-2