National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :	RADIOLOGICAL PRECISION LABS (INDIA) PVT. LTD., 1ST FLOOR, 18 BLUECHIP INDL ESTATE NO. 1, GOLANI INDL COMPLEX, WALIV, VASAI (EAST), VASAI, MAHARASHTRA, INDIA		
Accreditation Standard	ISO/IEC 17025:2017		
Certificate Number	TC-9574	**Page No**	1 of 9
Validity	03/06/2021 to 02/06/2023	**Last Amended on**	04/09/2022

S.No	Discipline / Group	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used
Site Facility				
1	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	CT Equipment	Measurement of Computed Tomography Dose Index (CTDI)	AERB/RF-MED/SC-03
2	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	CT Equipment	Accuracy of Gantry Tilt	AERB/RF-MED/SC-03
3	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	CT Equipment	High Contrast Resolution	AERB/RF-MED/SC-03
4	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	CT Equipment	Low contrast resolution	AERB/RF-MED/SC-03
5	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	CT Equipment	mA/mAs Linearity	AERB/RF-MED/SC-03
6	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	CT Equipment	Operating Potential	AERB/RF-MED/SC-03
7	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	CT Equipment	Output Consistency/Reproducibility of Output	AERB/RF-MED/SC-03
8	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	CT Equipment	Radiation Leakage Levels From X-ray Tube Housing	AERB/RF-MED/SC-03
9	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	CT Equipment	Radiation Profile Width/Slice Thickness	AERB/RF-MED/SC-03
10	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	CT Equipment	Radiation Protection Survey	AERB/RF-MED/SC-03
11	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	CT Equipment	Time	AERB/RF-MED/SC-03
12	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	CT Equipment	Total Filtration (TF)	AERB/RF-MED/SC-03
13	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	DENTAL MACHINE (Cone Beam Computed Tomography (CBCT))	CT Dose Index	AERB/RF-MED/SC-03

SCOPE OF ACCREDITATION

Laboratory Name : RADIOLOGICAL PRECISION LABS (INDIA) PVT. LTD., 1ST FLOOR, 18 BLUECHIP INDL ESTATE NO. 1, GOLANI INDL COMPLEX, WALIV, VASAI (EAST), VASAI, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number TC-9574 **Page No** 2 of 9

Validity 03/06/2021 to 02/06/2023 **Last Amended on** 04/09/2022

S.No	Discipline / Group	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used
14	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	DENTAL MACHINE (Cone Beam Computed Tomography (CBCT))	High Contrast Resolution	AERB/RF-MED/SC-03
15	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	DENTAL MACHINE (Cone Beam Computed Tomography (CBCT))	Linearity of mA/mAs Loading Station	AERB/RF-MED/SC-03
16	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	DENTAL MACHINE (Cone Beam Computed Tomography (CBCT))	Low Contrast Resolution	AERB/RF-MED/SC-03
17	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	DENTAL MACHINE (Cone Beam Computed Tomography (CBCT))	Operating Potential	AERB/RF-MED/SC-03
18	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	DENTAL MACHINE (Cone Beam Computed Tomography (CBCT))	Output Consistency/Reproducibility of Output	AERB/RF-MED/SC-03
19	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	DENTAL MACHINE (Cone Beam Computed Tomography (CBCT))	Radiation Leakage Levels from X-ray Tube Housing & Collimator	AERB/RF-MED/SC-03
20	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	DENTAL MACHINE (Cone Beam Computed Tomography (CBCT))	Radiation Protection Survey	AERB/RF-MED/SC-03
21	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	DENTAL MACHINE (Cone Beam Computed Tomography (CBCT))	Time	AERB/RF-MED/SC-03
22	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	DENTAL MACHINE (Cone Beam Computed Tomography (CBCT))	Timer Linearity	AERB/RF-MED/SC-03
23	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	DENTAL MACHINE (Cone Beam Computed Tomography (CBCT))	Total Filtration (TF)	AERB/RF-MED/SC-03
24	DIAGNOSTIC RADIOLOGY QA TESTING- INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	C- Arm & O- Arm	Radiation Leakage Levels from X-ray Tube Housing & Collimator	AERB/RF-MED/SC-03
25	DIAGNOSTIC RADIOLOGY QA TESTING- INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	C-Arm & O-Arm	Exposure Rate at Table Top	AERB/RF-MED/SC-03

SCOPE OF ACCREDITATION

Laboratory Name :	RADIOLOGICAL PRECISION LABS (INDIA) PVT. LTD., 1ST FLOOR, 18 BLUECHIP INDL ESTATE NO. 1, GOLANI INDL COMPLEX, WALIV, VASAI (EAST), VASAI, MAHARASHTRA, INDIA		
Accreditation Standard	ISO/IEC 17025:2017		
Certificate Number	TC-9574	**Page No**	3 of 9
Validity	03/06/2021 to 02/06/2023	**Last Amended on**	04/09/2022

S.No	Discipline / Group	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used
26	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	C-Arm & O-Arm	High Contrast Resolution	AERB/RF-MED/SC-03
27	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	C-Arm & O-Arm	Linearity of mA/mAs Loading Station	AERB/RF-MED/SC-03
28	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	C-Arm & O-Arm	Low Contrast Resolution	AERB/RF-MED/SC-03
29	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	C-Arm & O-Arm	Output Consistency/Reproducibility of Output	AERB/RF-MED/SC-03
30	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	C-Arm & O-Arm	Total Filtration (TF)	AERB/RF-MED/SC-03
31	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	C-Arm & O-Arm Machine	Operating Potential	AERB/RF-MED/SC-03
32	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	C-Arm & O-Arm Machine	Time	AERB/RF-MED/SC-03
33	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Interventional Radiology	Total Filtration (TF)	AERB/RF-MED/SC-03
34	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Interventional Radiology Machine	Exposure Rate at Table Top	AERB/RF-MED/SC-03

SCOPE OF ACCREDITATION

Laboratory Name :	RADIOLOGICAL PRECISION LABS (INDIA) PVT. LTD., 1ST FLOOR, 18 BLUECHIP INDL ESTATE NO. 1, GOLANI INDL COMPLEX, WALIV, VASAI (EAST), VASAI, MAHARASHTRA, INDIA
Accreditation Standard	ISO/IEC 17025:2017
Certificate Number	TC-9574
Page No	4 of 9
Validity	03/06/2021 to 02/06/2023
Last Amended on	04/09/2022

S.No	Discipline / Group	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used
35	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Interventional Radiology Machine	High Contrast Resolution	AERB/RF-MED/SC-03
36	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Interventional Radiology Machine	Linearity of mA/mAs Loading Station	AERB/RF-MED/SC-03
37	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Interventional Radiology Machine	Low Contrast Resolution	AERB/RF-MED/SC-03
38	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Interventional Radiology Machine	Operating Potential	AERB/RF-MED/SC-03
39	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Interventional Radiology Machine	Output Consistency/Reproducibility of Output	AERB/RF-MED/SC-03
40	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Interventional Radiology Machine	Radiation Leakage Levels from X-ray Tube Housing & Collimator	AERB/RF-MED/SC-03
41	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Interventional Radiology Machine	Radiation Protection Survey	AERB/RF-MED/SC-03
42	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Interventional Radiology Machine	Time	AERB/RF-MED/SC-03
43	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Radiography & Fluoroscopy Machine	Central Beam Alignment	AERB/RF-MED/SC-03

SCOPE OF ACCREDITATION

Laboratory Name :	RADIOLOGICAL PRECISION LABS (INDIA) PVT. LTD., 1ST FLOOR, 18 BLUECHIP INDL ESTATE NO. 1, GOLANI INDL COMPLEX, WALIV, VASAI (EAST), VASAI, MAHARASHTRA, INDIA		
Accreditation Standard	ISO/IEC 17025:2017		
Certificate Number	TC-9574	**Page No**	5 of 9
Validity	03/06/2021 to 02/06/2023	**Last Amended on**	04/09/2022

S.No	Discipline / Group	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used
44	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Radiography & Fluoroscopy Machine	Congruence of Radiation & Optical Field	AERB/RF-MED/SC-03
45	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Radiography & Fluoroscopy Machine	Effective Focal Spot Measurement	AERB/RF-MED/SC-03
46	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Radiography & Fluoroscopy Machine	Exposure Rate at Table Top	AERB/RF-MED/SC-03
47	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Radiography & Fluoroscopy Machine	High Contrast Resolution	AERB/RF-MED/SC-03
48	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Radiography & Fluoroscopy Machine	Linearity of mA/mAs Loading Station	AERB/RF-MED/SC-03
49	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Radiography & Fluoroscopy Machine	Low Contrast Resolution	AERB/RF-MED/SC-03
50	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Radiography & Fluoroscopy Machine	Operating Potential	AERB/RF-MED/SC-03
51	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Radiography & Fluoroscopy Machine	Output Consistency/Reproducibility of Output	AERB/RF-MED/SC-03
52	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Radiography & Fluoroscopy Machine	Radiation Leakage Levels from X-ray Tube Housing & Collimator	AERB/RF-MED/SC-03

SCOPE OF ACCREDITATION

Laboratory Name :	RADIOLOGICAL PRECISION LABS (INDIA) PVT. LTD., 1ST FLOOR, 18 BLUECHIP INDL ESTATE NO. 1, GOLANI INDL COMPLEX, WALIV, VASAI (EAST), VASAI, MAHARASHTRA, INDIA
Accreditation Standard	ISO/IEC 17025:2017
Certificate Number	TC-9574
Page No	6 of 9
Validity	03/06/2021 to 02/06/2023
Last Amended on	04/09/2022

S.No	Discipline / Group	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used
53	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Radiography & Fluoroscopy Machine	Radiation Protection Survey	AERB/RF-MED/SC-03
54	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Radiography & Fluoroscopy Machine	Time	AERB/RF-MED/SC-03
55	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Radiography & Fluoroscopy Machine	Timer Linearity	AERB/RF-MED/SC-03
56	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Radiography & Fluoroscopy Machine	Total Filtration (TF)	AERB/RF-MED/SC-03
57	DIAGNOSTIC RADIOLOGY QA TESTING-MAMMOGRAPHY	Mammography	Half Value Layer (HVL)	AERB/RF-MED/SC-03
58	DIAGNOSTIC RADIOLOGY QA TESTING-MAMMOGRAPHY	Mammography	Imaging Performance Evaluation - Masses	AERB/RF-MED/SC-03
59	DIAGNOSTIC RADIOLOGY QA TESTING-MAMMOGRAPHY	Mammography	Imaging Performance Evaluation - Micro Calcification	AERB/RF-MED/SC-03
60	DIAGNOSTIC RADIOLOGY QA TESTING-MAMMOGRAPHY	Mammography	Imaging Performance Evaluation -Fibers	AERB/RF-MED/SC-03
61	DIAGNOSTIC RADIOLOGY QA TESTING-MAMMOGRAPHY	Mammography	Linearity of mA/mAs Loading Station	AERB/RF-MED/SC-03
62	DIAGNOSTIC RADIOLOGY QA TESTING-MAMMOGRAPHY	Mammography	Operating Potential	AERB/RF-MED/SC-03
63	DIAGNOSTIC RADIOLOGY QA TESTING-MAMMOGRAPHY	Mammography	Output Consistency/Reproducibility of Output	AERB/RF-MED/SC-03

SCOPE OF ACCREDITATION

Laboratory Name :	RADIOLOGICAL PRECISION LABS (INDIA) PVT. LTD., 1ST FLOOR, 18 BLUECHIP INDL ESTATE NO. 1, GOLANI INDL COMPLEX, WALIV, VASAI (EAST), VASAI, MAHARASHTRA, INDIA		
Accreditation Standard	ISO/IEC 17025:2017		
Certificate Number	TC-9574	**Page No**	7 of 9
Validity	03/06/2021 to 02/06/2023	**Last Amended on**	04/09/2022

S.No	Discipline / Group	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used
64	DIAGNOSTIC RADIOLOGY QA TESTING-MAMMOGRAPHY	Mammography	Radiation Leakage Levels from X-ray Tube Housing & Collimator	AERB/RF-MED/SC-03
65	DIAGNOSTIC RADIOLOGY QA TESTING-MAMMOGRAPHY	Mammography	Radiation Protection Survey	AERB/RF-MED/SC-03
66	DIAGNOSTIC RADIOLOGY QA TESTING-MAMMOGRAPHY	Mammography	Time	AERB/RF-MED/SC-03
67	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	DENTAL MACHINE (Intra Oral Peri-apical radiograph (IOPA))	Resolution of Image Sensor (in lp/mm)	AERB/RF-MED/SC-03
68	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	DENTAL MACHINE (Intra Oral Peri-apical radiograph (IOPA), Ortho Pan tomography (OPG))	Radiation Protection Survey	AERB/RF-MED/SC-03
69	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	DENTAL MACHINE (Intra Oral Peri-apical radiograph (IOPA), Ortho Pan tomography (OPG), Dental Hand Held)	Measurement of Linearity of mA/mAs Loading Station	AERB/RF-MED/SC-03
70	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	DENTAL MACHINE (Intra Oral Peri-apical radiograph (IOPA), Ortho Pan tomography (OPG), Dental Hand Held)	Measurement of Operating Potential	AERB/RF-MED/SC-03
71	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	DENTAL MACHINE (Intra Oral Peri-apical radiograph (IOPA), Ortho Pan tomography (OPG), Dental Hand Held)	Output Consistency/Reproducibility of Output	AERB/RF-MED/SC-03
72	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	DENTAL MACHINE (Intra Oral Peri-apical radiograph (IOPA), Ortho Pan tomography (OPG), Dental Hand Held)	Radiation Leakage Levels from X-ray Tube Housing & Collimator	AERB/RF-MED/SC-03
73	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	DENTAL MACHINE (Intra Oral Peri-apical radiograph (IOPA), Ortho Pan tomography (OPG), Dental Hand Held)	Timer Accuracy	AERB/RF-MED/SC-03
74	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	DENTAL MACHINE (Intra Oral Peri-apical radiograph (IOPA), Ortho Pan tomography (OPG), Dental Hand Held)	Timer Linearity	AERB/RF-MED/SC-03

SCOPE OF ACCREDITATION

Laboratory Name : RADIOLOGICAL PRECISION LABS (INDIA) PVT. LTD., 1ST FLOOR, 18 BLUECHIP INDL ESTATE NO. 1, GOLANI INDL COMPLEX, WALIV, VASAI (EAST), VASAI, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number TC-9574 **Page No** 8 of 9

Validity 03/06/2021 to 02/06/2023 **Last Amended on** 04/09/2022

S.No	Discipline / Group	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used
75	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	DENTAL MACHINE (Intra Oral Peri-apical radiograph (IOPA), Ortho Pan tomography (OPG), Dental Hand Held)	Total Filtration (TF)	AERB/RF-MED/SC-03
76	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	DEXA (Dual energy X-ray Absorptiometry)	Radiation Leakage Levels from X-ray Tube Housing & Collimator	AERB/RF-MED/SC-03
77	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	DEXA (Dual energy X-ray Absorptiometry)	Radiation Protection Survey	AERB/RF-MED/SC-03
78	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	X RAY (FIXED, PORTABLE, MOBILE) Machine	Central Beam Alignment	AERB/RF-MED/SC-03
79	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	X RAY (FIXED, PORTABLE, MOBILE) Machine	Congruence of Radiation & Optical Field	AERB/RF-MED/SC-03
80	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	X RAY (FIXED, PORTABLE, MOBILE) Machine	Effective Focal Spot Measurement	AERB/RF-MED/SC-03
81	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	X RAY (FIXED, PORTABLE, MOBILE) Machine	Measurement of Linearity of mA/mAs Loading Station	AERB/RF-MED/SC-03
82	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	X RAY (FIXED, PORTABLE, MOBILE) Machine	Measurement of Operating Potential	AERB/RF-MED/SC-03
83	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	X RAY (FIXED, PORTABLE, MOBILE) Machine	Output Consistency/Reproducibility of Output	AERB/RF-MED/SC-03
84	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	X RAY (FIXED, PORTABLE, MOBILE) Machine	Radiation Leakage Levels from X-ray Tube Housing & Collimator	AERB/RF-MED/SC-03
85	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	X RAY (FIXED, PORTABLE, MOBILE) Machine	Radiation Protection Survey	AERB/RF-MED/SC-03

SCOPE OF ACCREDITATION

Laboratory Name :	RADIOLOGICAL PRECISION LABS (INDIA) PVT. LTD., 1ST FLOOR, 18 BLUECHIP INDL ESTATE NO. 1, GOLANI INDL COMPLEX, WALIV, VASAI (EAST), VASAI, MAHARASHTRA, INDIA		
Accreditation Standard	ISO/IEC 17025:2017		
Certificate Number	TC-9574	**Page No**	9 of 9
Validity	03/06/2021 to 02/06/2023	**Last Amended on**	04/09/2022

S.No	Discipline / Group	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used
86	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	X RAY (FIXED, PORTABLE, MOBILE) Machine	Timer Accuracy	AERB/RF-MED/SC-03
87	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	X RAY (FIXED, PORTABLE, MOBILE) Machine	Timer Linearity	AERB/RF-MED/SC-03
88	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	X RAY (FIXED, PORTABLE, MOBILE) Machine	Total Filtration (TF)	AERB/RF-MED/SC-03