National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name : CHEMICAL LABORATORY - UKAI THERMAL POWER STATION, UKAI DAM, TALUKA SONGADH, UKAI, TAPI, GUJARAT, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number TC-11653 **Page No** 1 of 1

Validity 16/05/2023 to 15/05/2025 **Last Amended on** -

S.No	Discipline / Group	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used
Permanent Facility				
1	CHEMICAL- SOLID FUELS	Coal	Ash (Air dry basis)	IS 1350 (Part 1)
2	CHEMICAL- SOLID FUELS	Coal	Gross Calorific Value (Air dry basis)	IS 1350 (Part 2)
3	CHEMICAL- SOLID FUELS	Coal	Inherent Moisture (Air dry basis)	IS 1350 (Part 1)
4	CHEMICAL- SOLID FUELS	Coal	Sulphur (Air dry basis)	(Withdrawn) ASTM D3177
5	CHEMICAL- SOLID FUELS	Coal	Total Moisture	IS 1350 (Part 1)
6	CHEMICAL- SOLID FUELS	Coal	Volatile Matter (Air dry basis)	IS 1350 (Part 1)