National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :	DISTRICT LABORATORY, GUJARAT WATER SUPPLY & SEWERAGE BOARD, PORBANDAR, JAL BHAVAN, PORBANDAR, GUJARAT, INDIA		
Accreditation Standard	ISO/IEC 17025:2017		
Certificate Number	TC-9701	**Page No**	1 of 1
Validity	26/07/2021 to 25/07/2023	**Last Amended on**	11/01/2023

S.No	Discipline / Group	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used
Permanent Facility				
1	CHEMICAL- WATER	Drinking Water	Calcium (as Ca)	APHA 23rd Edition -3500-Ca Method B (EDTA Method)
2	CHEMICAL- WATER	Drinking Water	Chloride (as Cl)	APHA 23rd Edition -4500-Cl Method B (Argentometric Method)
3	CHEMICAL- WATER	Drinking Water	Colour	APHA 23rd Edition -2120-Colour Method B (Platinum Cobalt Visual Comparison)
4	CHEMICAL- WATER	Drinking Water	Fluoride (as F)	APHA 23rd Edition -4500-F Method C (Electrode Method)
5	CHEMICAL- WATER	Drinking Water	Magnesium (as Mg)	APHA 23rd Edition -3500-Mg Method B (By Calculation)
6	CHEMICAL- WATER	Drinking Water	Nitrate (as NO3)	APHA 23rd Edition -4500-NO3 Method B (UV Screening Method)
7	CHEMICAL- WATER	Drinking Water	Odour	IS 3025 (Part 5)
8	CHEMICAL- WATER	Drinking Water	pH	APHA 23rd Edition -4500-H+ Method B
9	CHEMICAL- WATER	Drinking Water	Sulphate (as SO4)	APHA 23rd Edition -4500-SO4 Method E (Turbidimetric Method)
10	CHEMICAL- WATER	Drinking Water	Taste	IS 3025 (Part 8)
11	CHEMICAL- WATER	Drinking Water	Total Alkalinity (as CaCO3)	APHA 23rd Edition -2320 Alkalinity Method B (Titration Method)
12	CHEMICAL- WATER	Drinking Water	Total Dissolved Solids	APHA 23rd Edition -2540 TDS Method C (Dried at 180 Degree Celsius)
13	CHEMICAL- WATER	Drinking Water	Total Hardness (as CaCO3)	APHA 23rd Edition -2340-Hardness Method C (EDTA Method)
14	CHEMICAL- WATER	Drinking Water	Turbidity	APHA 23rd Edition -2130 Turbidity Method B