National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name : SCHNELL MEDICAL SERVICES, 215, GOPI CINE MALL, NSS ROAD, DOMBIVLI WEST, THANE, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number TC-9469 **Page No** 1 of 9

Validity 24/04/2023 to 23/04/2025 **Last Amended on** -

S.No	Discipline / Group	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used
Site Facility				
1	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	Computed Tomography (CT)	Alignment of table to gantry	AERB/RF-MED/SC-03
2	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	Computed Tomography (CT)	Gantry tilt	AERB/RF-MED/SC-03
3	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	Computed Tomography (CT)	High Contrast Resolution	AERB/RF-MED/SC-03
4	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	Computed Tomography (CT)	Low Contrast Resolution	AERB/RF-MED/SC-03
5	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	Computed Tomography (CT)	mA/mAS Linearity (CoL)	AERB/RF-MED/SC-03
6	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	Computed Tomography (CT)	Operating Potential	AERB/RF-MED/SC-03
7	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	Computed Tomography (CT)	Radiation dose test CTDIw	AERB/RF-MED/SC-03
8	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	Computed Tomography (CT)	Radiation leakage level from tube housing	AERB/RF-MED/SC-03
9	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	Computed Tomography (CT)	Radiation protection survey	AERB/RF-MED/SC-03
10	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	Computed Tomography (CT)	Reproducibility of output (CoV)	AERB/RF-MED/SC-03
11	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	Computed Tomography (CT)	Slice Thickness	AERB/RF-MED/SC-03
12	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	Computed Tomography (CT)	Table Indexing	AERB/RF-MED/SC-03
13	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	Computed Tomography (CT)	Time	AERB/RF-MED/SC-03

SCOPE OF ACCREDITATION

Laboratory Name : SCHNELL MEDICAL SERVICES, 215, GOPI CINE MALL, NSS ROAD, DOMBIVLI WEST, THANE, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number TC-9469 **Page No** 2 of 9

Validity 24/04/2023 to 23/04/2025 **Last Amended on** -

S.No	Discipline / Group	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used
14	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	Computed Tomography (CT)	Total Filtration	AERB/RF-MED/SC-03
15	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	Dental Cone Beam CT	CTDIw	AERB/RF-MED/SC-03
16	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	Dental Cone Beam CT	High contrast resolution	AERB/RF-MED/SC-03
17	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	Dental Cone Beam CT	Low contrast resolution	AERB/RF-MED/SC-03
18	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	Dental Cone Beam CT	mA/mAs linearity (CoL)	AERB/RF-MED/SC-03
19	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	Dental Cone Beam CT	Operating Potential	AERB/RF-MED/SC-03
20	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	Dental Cone Beam CT	Radiation Protection Survey	AERB/RF-MED/SC-03
21	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	Dental Cone Beam CT	Reproducibility of output(CoV)	AERB/RF-MED/SC-03
22	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	Dental Cone Beam CT	Time	AERB/RF-MED/SC-03
23	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	Dental Cone Beam CT	Total filtration (HVL)	AERB/RF-MED/SC-03
24	DIAGNOSTIC RADIOLOGY QA TESTING- COMPUTED TOMOGRAPHY (CT)	Dental Cone Beam CT	Tube Housing Leakage	AERB/RF-MED/SC-03
25	DIAGNOSTIC RADIOLOGY QA TESTING- INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	C-arm and O-arm	Collimator leakage	AERB/RF-MED/SC-03

SCOPE OF ACCREDITATION

Laboratory Name : SCHNELL MEDICAL SERVICES, 215, GOPI CINE MALL, NSS ROAD, DOMBIVLI WEST, THANE, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number TC-9469 **Page No** 3 of 9

Validity 24/04/2023 to 23/04/2025 **Last Amended on** -

S.No	Discipline / Group	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used
26	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	C-arm and O-arm	Consistency of Radiation Output	AERB/RF-MED/SC-03
27	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	C-arm and O-arm	Exposure Rate at Tabletop	AERB/RF-MED/SC-03
28	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	C-arm and O-arm	High contrast resolution	AERB/RF-MED/SC-03
29	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	C-Arm and O-Arm	Irradiation Time	AERB/RF-MED/SC-03
30	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	C-Arm and O-Arm	Linearity of mA/mAs loading Stations	AERB/RF-MED/SC-03
31	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	C-arm and O-arm	Low contrast resolution	AERB/RF-MED/SC-03
32	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	C-Arm and O-Arm	Operating Potential	AERB/RF-MED/SC-03
33	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	C-Arm and O-Arm	Total filtration (HVL)	AERB/RF-MED/SC-03
34	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	C-arm and O-arm	Tube housing leakage	AERB/RF-MED/SC-03

SCOPE OF ACCREDITATION

Laboratory Name : SCHNELL MEDICAL SERVICES, 215, GOPI CINE MALL, NSS ROAD, DOMBIVLI WEST, THANE, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number TC-9469 **Page No** 4 of 9

Validity 24/04/2023 to 23/04/2025 **Last Amended on** -

S.No	Discipline / Group	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used
35	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Cathlab	Consistency of radiation output	AERB/RF-MED/SC-03
36	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Cathlab	Exposure Rate at Table Top	AERB/RF-MED/SC-03
37	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Cathlab	High Contrast Resolution	AERB/RF-MED/SC-03
38	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Cathlab	Irradiation time	AERB/RF-MED/SC-03
39	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Cathlab	Linerity of mA/mAs loading stations	AERB/RF-MED/SC-03
40	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Cathlab	Low Contrast Resolution	AERB/RF-MED/SC-03
41	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Cathlab	Operating Potential	AERB/RF-MED/SC-03
42	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Cathlab	Radiation leakage level from tube housing and Collimator	AERB/RF-MED/SC-03
43	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Cathlab	Radiation Protection Survey	AERB/RF-MED/SC-03

SCOPE OF ACCREDITATION

Laboratory Name : SCHNELL MEDICAL SERVICES, 215, GOPI CINE MALL, NSS ROAD, DOMBIVLI WEST, THANE, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number TC-9469 **Page No** 5 of 9

Validity 24/04/2023 to 23/04/2025 **Last Amended on** -

S.No	Discipline / Group	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used
44	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Cathlab	Total Filtration	AERB/RF-MED/SC-03
45	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Radiography & Fluoroscopy	Central Beam alignment	AERB/RF-MED/SC-03
46	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Radiography & Fluoroscopy	Congruence of radiation and optical field	AERB/RF-MED/SC-03
47	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Radiography & Fluoroscopy	Effective focal spot measurement	AERB/RF-MED/SC-03
48	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Radiography & Fluoroscopy	Exposure Rate at Tabletop	AERB/RF-MED/SC-03
49	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Radiography & Fluoroscopy	High Contrast Resolution	AERB/RF-MED/SC-03
50	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Radiography & Fluoroscopy	Linearity of mA/mAs loading Stations	AERB/RF-MED/SC-03
51	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Radiography & Fluoroscopy	Low contrast resolution	AERB/RF-MED/SC-03
52	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Radiography & Fluoroscopy	Operating Potential	AERB/RF-MED/SC-03

SCOPE OF ACCREDITATION

Laboratory Name : SCHNELL MEDICAL SERVICES, 215, GOPI CINE MALL, NSS ROAD, DOMBIVLI WEST, THANE, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number TC-9469 **Page No** 6 of 9

Validity 24/04/2023 to 23/04/2025 **Last Amended on** -

S.No	Discipline / Group	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used
53	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Radiography & Fluoroscopy	Radiation Protection Survey	AERB/RF-MED/SC-03
54	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Radiography & Fluoroscopy	Reproducibility of Output (CoV)	AERB/RF-MED/SC-03
55	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Radiography & Fluoroscopy	Time	AERB/RF-MED/SC-03
56	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Radiography & Fluoroscopy	Total filtration (HVL)	AERB/RF-MED/SC-03
57	DIAGNOSTIC RADIOLOGY QA TESTING-INTERVENTIONAL RADIOLOGY & FLUOROSCOPY	Radiography & Fluoroscopy	Tube housing leakage	AERB/RF-MED/SC-03
58	DIAGNOSTIC RADIOLOGY QA TESTING-MAMMOGRAPHY	Mammography	mA/mAS Linearity	AERB/RF-MED/SC-03
59	DIAGNOSTIC RADIOLOGY QA TESTING-MAMMOGRAPHY	Mammography	Operating Potential	AERB/RF-MED/SC-03
60	DIAGNOSTIC RADIOLOGY QA TESTING-MAMMOGRAPHY	Mammography	Performance of imaging phantom	AERB/RF-MED/SC-03
61	DIAGNOSTIC RADIOLOGY QA TESTING-MAMMOGRAPHY	Mammography	Radiation Leakage level from X Ray tube housing	AERB/RF-MED/SC-03
62	DIAGNOSTIC RADIOLOGY QA TESTING-MAMMOGRAPHY	Mammography	Radiation protection survey	AERB/RF-MED/SC-03
63	DIAGNOSTIC RADIOLOGY QA TESTING-MAMMOGRAPHY	Mammography	Reproducibility of output	AERB/RF-MED/SC-03

SCOPE OF ACCREDITATION

Laboratory Name : SCHNELL MEDICAL SERVICES, 215, GOPI CINE MALL, NSS ROAD, DOMBIVLI WEST, THANE, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number TC-9469 **Page No** 7 of 9

Validity 24/04/2023 to 23/04/2025 **Last Amended on** -

S.No	Discipline / Group	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used
64	DIAGNOSTIC RADIOLOGY QA TESTING-MAMMOGRAPHY	Mammography	Time	AERB/RF-MED/SC-03
65	DIAGNOSTIC RADIOLOGY QA TESTING-MAMMOGRAPHY	Mammography	Total Filtration/HVL	AERB/RF-MED/SC-03
66	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	Bone Mineral Densitometry/ DEXA	Operating Potential	AERB/RF-MED/SC-03
67	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	Bone Mineral Densitometry/DEXA	Linearity of mA/mAs loading Stations	AERB/RF-MED/SC-03
68	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	Bone Mineral Densitometry/DEXA	Radiation Protection Survey	AERB/RF-MED/SC-03
69	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	Bone Mineral Densitometry/DEXA	Reproducibility of output(CoV)	AERB/RF-MED/SC-03
70	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	Bone Mineral Densitometry/DEXA	Time	AERB/RF-MED/SC-03
71	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	Bone Mineral Densitometry/DEXA	Total filtration (HVL)	AERB/RF-MED/SC-03
72	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	Bone Mineral Densitometry/DEXA	Tube Housing Leakage	AERB/RF-MED/SC-03
73	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	Dental (OPG and Intra Oral)	High contrast resolution	AERB/RF-MED/SC-03
74	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	Dental (OPG and Intra Oral)	Low contrast resolution	AERB/RF-MED/SC-03

SCOPE OF ACCREDITATION

Laboratory Name : SCHNELL MEDICAL SERVICES, 215, GOPI CINE MALL, NSS ROAD, DOMBIVLI WEST, THANE, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number TC-9469 **Page No** 8 of 9

Validity 24/04/2023 to 23/04/2025 **Last Amended on** -

S.No	Discipline / Group	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used
75	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	Dental (OPG and Intra Oral)	mA/mAs linearity (CoL)	AERB/RF-MED/SC-03
76	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	Dental (OPG and Intra Oral)	Operating Potential	AERB/RF-MED/SC-03
77	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	Dental (OPG and Intra Oral)	Radiation Protection Survey	AERB/RF-MED/SC-03
78	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	Dental (OPG and Intra Oral)	Reproducibility of output(CoV)	AERB/RF-MED/SC-03
79	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	Dental (OPG and Intra Oral)	Time	AERB/RF-MED/SC-03
80	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	Dental (OPG and Intra Oral)	Total filtration (HVL)	AERB/RF-MED/SC-03
81	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	Dental (OPG and Intra Oral)	Tube Housing Leakage	AERB/RF-MED/SC-03
82	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	Radiography X-ray Machines	Central Beam Alignment	AERB/RF-MED/SC-03
83	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	Radiography X-ray Machines	Congruence of radiation and optical field	AERB/RF-MED/SC-03
84	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	Radiography X-ray Machines	Consistancy of radiation output	AERB/RF-MED/SC-03
85	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	Radiography X-ray Machines	Effective focal spot	AERB/RF-MED/SC-03

SCOPE OF ACCREDITATION

Laboratory Name : SCHNELL MEDICAL SERVICES, 215, GOPI CINE MALL, NSS ROAD, DOMBIVLI WEST, THANE, MAHARASHTRA, INDIA

Accreditation Standard ISO/IEC 17025:2017

Certificate Number TC-9469 **Page No** 9 of 9

Validity 24/04/2023 to 23/04/2025 **Last Amended on** -

S.No	Discipline / Group	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used
86	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	Radiography X-ray Machines	Irradiation Time	AERB/RF-MED/SC-03
87	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	Radiography X-ray Machines	Linerity of mA/mAs loading stations	AERB/RF-MED/SC-03
88	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	Radiography X-ray Machines	Operating Potential	AERB/RF-MED/SC-03
89	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	Radiography X-ray Machines	Radiation leakage level at 1 m from tube housing and Collimator	AERB/RF-MED/SC-03
90	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	Radiography X-ray Machines	Radiation protection survey	AERB/RF-MED/SC-03
91	DIAGNOSTIC RADIOLOGY QA TESTING-RADIOGRAPHY (X-RAY - FIXED PORTABLE, MOBILE)	Radiography X-ray Machines	Total filtration	AERB/RF-MED/SC-03